

Corrigendum

There is **change in Specification for Atomic Absorption Spectrophotometer** for Annexure -II A.

Sl. No	Criteria	Tender Specification	To be Amended as
3	Wavelength Range	175 - 915 nm Diffraction grating min:1800/mm lines and Reciprocal linear Dispersion greater than 0.45nm/mm @200 nm, focal length 300mm or more	185 - 900 nm Diffraction grating min:1800/mm lines and Reciprocal linear Dispersion greater than 0.45nm/mm @200 nm, focal length 300mm or more
10	Detector	PMT (Photomultiplier)	PMT (Photomultiplier)/ Solid state detector.