
This tender floated in GeM portal. Bidder has to submit their offer through GeM only.








आईआरईएल (इंडिया) लिमिटेड IREL (India) Limited

(पूर्वमेंइंडियनरेअरअर्थ्सलिमिटेड Formerly Indian Rare Earths Ltd.) रेअरअर्थ्सप्रभाग Rare Earths Division,





IREL/RED/P/24-25/31653/

निविदा आमंत्रण सूचना NOTICE INVITING TO TENDER

SCHEDULE OF TENDER (SOT)

a.	Description of supply	Supply, installation, commissioning, testing &
		training of HP Ge detector based gamma
		spectrometer and accessories
b.	Mode of tender	Open tender invited in two parts.
		Part-I: Techno-Commercial Bid
		Part - II: Price Bid.
С	Pre-Bid meeting	04-04-2025, 11.00Hrs, at IREL, RED,
		Udyogamandal
c.	Estimated Value	Rs.1,93,99,200/- (all inclusive)
d.	Earnest Money Deposit (EMD)	Rs. 3,28,800/-
e.	Validity of Tender	90 days from the date of opening of Techno-
		Commercial bid.
f.	Contact details of tender inviting	Shri.V A Anil Kumar- CM-Technical (Purchase)
	authority	IREL (India) Limited, R E Division,
		Udyogamandal – 683 501, KERALA
		E-mail: <u>purchase-red@irel.co.in</u>
		Ph. No. 0484-2545199

NOTE

EMD EXEMPTION: For this tender, Under MSE category, only manufacturers for goods are eligible for exemption from EMD. Traders are excluded from the purview of this Policy. The bidder seeking EMD exemption, must submit the valid supporting document for the relevant category along with Udyam Certificate otherwise offer will be rejected without any further clarification.

BIDDING SYSTEMS

This is a two-part tender. The tenderers should bifurcate their quotations in two parts (Part-I & II).

PART-I "Techno-commercial bid":

It contains the eligibility, technical quality and performance aspects, EMD, commercial terms and conditions and documents sought in the tender, **except the price and relevant financial details.**

PART-II "Financial bid":

It contains the price quotation along with other financial details.

PART-I shall be opened first and evaluated by a committee constituted by competent authority.

PART-II: the financial bid of Techno-commercially eligible bidders shall be opened on recommendation of above committee on a notified date and time after sending advance notice to all qualified bidders for being present.

CONTACT PERSONS OF TENDERING AUTHORITY

Purpose	Name	E-mail ID	Contact No.
For Bid/tender	Shri. V A Anil Kumar	purchase-red@irel.co.in	0484-2545199
related query	CM-Technical (Purchase)		9443482644
For Technical	Smt. Neetha Rajesh	qc-red@irel.co.om	91 9004770584
Specification / Scope	Manager-Technical (QC)		
of work related			
query			
For	Shri. Sanjaykumar Vind	stores-red@irel.co.in	91 8301997625
Consignment/Goods	SM-Technical (Stores)		
Transportation			
related query			
For Payment/refund	Smt. P Kamalam	finance-red@irel.co.in	91 8593994144
related query	DGM - Finance (A & E)		

EARNEST MONEY DEPOSIT (EMD)/ BID SECURITY

- 1.0 Earnest Money Deposit (EMD) is a deposit received from the tenderers in token of their earnestness in submitting their offer to undertake the supplies/works/services/consultancy contracts and conclude a contract if entrusted to them on the basis of their tender. The offers received from tenderers without EMD shall be summarily rejected except where an exemption is provided in the tender.
- 2.0 EMD is to be remitted by way of 'Insurance Surety Bonds' or 'account payee demand draft' or 'fixed deposit receipt' or 'bankers cheque' or 'Bank Guarantee from any scheduled commercial Banks' or "online payment' in favour of IREL. In case of online payment, bidders should submit the UTR no. enabling verification of receipt of the amount.
- 3.0 Government Body/Public Sector Undertakings may be exempted from payment of EMD with the approval of Competent Authority.
- 4.0 EMD amount in rupee value is to be mentioned as a fixed amount in the tender and not as a percentage of the estimated cost and no interest is payable on the EMD.
- 5.0 EMD shall be exempted to MSEs and Start-ups as per prevailing guidelines in this regard. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD.
- 6.0 EMD is liable to be forfeited if:
 - a) The tenderer changes the terms and conditions or prices or withdraw his quotation subsequent to the date of opening/ The tenderer impairs or derogates from the tender in any respect within the period of validity of the tender
 - b) The tenderer fails to accept the order when placed or fails to commence supplies/works/services after accepting the order.
 - c) In case bidder submits false/fabricated documents. d) In case bidder fails to submit SD as stipulated in the tender.
- 7.0 EMD may be adjusted against security deposit / performance security of the successful bidder. EMD of the unsuccessful bidders should be returned to them at the earliest after expiry of the final bid validity period and latest by the 30th day after the award of the contract. Bid security should be refunded to the successful bidder on receipt of performance security. However, in case of two stage bidding/ 2or 3 envelope bidding, EMD of unsuccessful bidders during first stage i.e. technical evaluation etc. should be returned within 30 days of declaration of result of first stage i.e. technical evaluation etc.
- 8.0 In case where the EMD is provided in form of BG in the prescribed format to be attached with the tender, the BG shall be obtained from a scheduled commercial Bank with validity of 45 days beyond final bid validity period.

GUIDELINES FOR PROCUREMENT FROM MSEs

- 1.0 Procurement from Micro, Small and Medium Enterprises (MSEs):
 - The Procurement Policy for Micro and Small Enterprises, 2012 [as amended time to time] has been notified by the Government in exercise of the powers conferred in Section 11 of the Micro, Small and Medium Enterprises Development (MSMED) Act, 2006. In line with Government Policy, the prevailing guidelines in this regard as issued from time to time is to be strictly adhered to.
- 2.0 As per the Government notified Public Procurement Policy for Micro & Small Enterprises (MSE), it is mandated that 25% of procurement of annual requirement of goods and services will be sourced from the Micro and Small Enterprises.
- 3.0 In tender, participating Micro and Small Enterprises (MSE) quoting price within price band of L1+15 (fifteen) per cent shall also be allowed to supply a portion of requirement by bringing down their price to L1 price in a situation where L1 price is from someone other than a MSE and such MSE shall be allowed to supply up to 25(twenty-five) per cent of total tendered value. The 25(twenty five) per cent quantity is to be distributed proportionately among these bidders, in case there are more than one MSEs within such price band.
- 4.0 Within this 25% (Twenty Five Percent) quantity, a purchase preference of four (4) per cent s reserved for MSEs owned by Scheduled Caste (SC)/ Scheduled Tribe (ST) entrepreneurs and three (3) percent is reserved for MSEs owned by women entrepreneur (if they participate in the tender process and match the L1 price). However, in event of failure of such MSEs to participate in tender process or meet tender requirements and L1 price, four percent sub-target for procurement earmarked for MSEs owned by SC/ST entrepreneurs and three (3) percent earmarked to women entrepreneur will be met from other MSEs.
- 5.0 MSEs would be treated as owned by SC/ ST entrepreneurs:
 - a) In case of proprietary MSE, proprietor(s) shall be SC /ST;
 - b) In case of partnership MSE, the SC/ ST partners shall be holding at least 51% (fifty-one percent) shares in the unit:
 - c) In case of Private Limited Companies, at least 51% (fifty-one percent) share shall be held by SC/ ST promoters.
- 6.0 If subcontract is given to MSEs, it will be considered as procurement from MSEs.
- 7.0 In case of tender item cannot be split or divided, etc. the MSE quoting a price within the band L1+15% may be awarded for full/ complete supply of total tendered value to MSE, considering the spirit of the Policy for enhancing Govt. Procurement from MSEs.
- 8.0 Where any Aggregator has been appointed by the Ministry of MSME, themselves quote on behalf of some MSE units, such offers will be considered as offers from MSE units and all such facilities would be extended to these also.
- 9.0 Traders/ distributors/ sole agent/ Works Contract are excluded from the purview of the policy.
- 10.0 To reduce transaction cost of doing business, Micro and Small Enterprises is to be facilitated by providing them tender sets free of cost, exempting Micro and Small Enterprises from payment of Earnest Money (EMD), adopting e-procurement to bring in transparency in tendering process. However, exemption from paying Performance Security / Performance Bank Guarantee is not covered under the Public Procurement policy.
- MSEs may also be given relaxation in prior turnover and prior experience criteria during the tender process, subject to meeting of quality and technical specifications. However, there may be circumstances (like procurement of items related to public safety, health, critical security operations and equipment, etc.) where the vendor having prior experience may be preferred rather than giving orders to new entities.
- 12.0 Payment to the MSME vendors is to be made within 30 days from the date of receipt and acceptance of goods/ services.
- 13.0 In case Bidder is a Micro or Small Enterprise under the Micro, Small and Medium Enterprises Development Act, 2006, the Bidder is to be asked for submitting the following along with tender:
- Ministry of MSME has notified criteria for classifying the enterprises as Micro, Small and Medium Enterprises as combination of investment in Plant & machinery or Equipment and Turnover. Mandatory registration in Udyam Registration Portal for all new and existing MSME has been prescribed. Accordingly, Micro and Small Enterprises (MSEs) shall be required to submit Udyam Registration Certificate for availing benefit.

15.0 RXIL (TReDS) PLATFORM

IREL is registered with RXIL (TReDS) platform. MSE bidders are requested to get registered with RXIL (TReDS) platform to avail the facility as per the GOI guidelines.

IREL (India) Limited is onboard with TReDS platform of M/s RXIL for facilitating Bill Discounting for MSME's.

As per DPE/7(4)/2007-Fin dt 21/08/2020, it is mandatory to get all the MSME vendors registered on the platform.

Vendors can get themselves registered at: https://onboarding.rxil.in/customerapp/home.

The registration fees of MSMEs on TReDS Platform is Free of Cost as per the new guidelines provided by SIDBI.

For any registration queries, vendors may please contact,

RXIL Relationship manager Mr. Satyajeet Jathar: +91 99201 00784/+91 9004100784 email: satyajeet.jathar@rxil.in

RXIL Relationship manager Mr Kirti musale: +91 90048 17501 email: kirti.musale@rxil.in

IREL (India) Limited Unit Administrator Mr. V A Anil Kumar: +91 9443482644

email: purchase-red@irel.co.in

IREL (India) Limited Nodal Officer Mr.

K.V.Ramakrishna: +918104997177 email: kvramakrishna@irel.co.in

SECURITY DEPOSIT (SD), PERFORMANCE BANK GUARANTEE & RETENTION MONEY

- 1.0 SECURITY DEPOSIT (PERFORMANCE SECURITY) & PERFORMANCE BANK GUARANTEE
- 1.1 Security deposit (SD) shall be uniformly levied @ 5% for the supply part (excluding Taxes) towards satisfactory completion of the order/works as under:
 - a) For supply & service contract valued more than Rs.5 lakhs. Performance Security is to be furnished by a specified date (generally 14 (fourteen) days after notification of the award) and it should remain valid for a period of 60 (sixty) days beyond the date of completion of all contractual obligations of the supplier/ contractor, including warranty obligations/ defect liability period (DLP).
- 1.2 Security deposit (SD) shall be uniformly levied @ 5% for the AMC part (excluding Taxes) after comprehensive warranty period & before starting of AMC. SD amount for AMC value will be adjusted from SD amount of the Supply part after comprehensive warranty period.
- 1.3 Submission of SD/PBG, if called for in the tender can also be paid through Insurance Surety Bonds' or 'account payee demand draft' or 'fixed deposit receipt from a Scheduled Commercial bank' or 'Bank Guarantee issued/ confirmed from any Scheduled Commercial Banks in India' or online payment to be confirmed sharing Unique Transaction Reference (UTR) to the tender inviting authority as called for in tender.
- 1.4 Bank Guarantees towards Bid Security/ Security Deposit/ PBG issued by Nationalised Banks/ Scheduled Commercial Banks (other than Cooperative Banks) only acceptable.
- 1.5 EMD/ Bid Security may be adjusted towards SD. However, if EMD is submitted in the form of Bank Guarantee, fresh Bank Guarantee in the prescribed format or demand draft/ bankers cheque is to be submitted towards SD.
- 1.6 The SD shall not bear any interest and is liable to be forfeited for unsatisfactory completion or on abandonment of the supply/ work order.

2.0 **RETENTION MONEY**

In contract, where payment is made on progressive billing of supply made/ work executed, 5% of the bill value is to be retained at the time of making payment towards rectification/defective work/supply made as retention money.

3.0 **SECURITY DEPOSIT & RETENTION MONEY**

The total SD (for supply part) and retention money together towards performance guarantee shall not exceed 10% of contract value. The performance security should be refunded to the supplier/ contractor without interest, after he duly performs and completes the contract in all respects but not later than 60(sixty) days of completion of all such obligations including the warranty under the contract/ 365 days beyond DLP.

4.0 **REFUND OF SD & RETENTION MONEY**

- a) Before releasing SD or retention money in respect of supplies/works, a "No Dues Certificate" shall be issued by EIC/ OIC duly countersigned by Head of the department after ensuring that no amounts are recoverable from the supplier/ contractor.
- b) EIC/OIC shall recommend release of SD and retention money after compliance by the contractor towards guarantee/ warranty/ performance guarantee & other related clauses as stipulated in the purchase/work order and on submission of formal claim by supplier/contractor.
- c) On receipt of "no dues certificate" from EIC/OIC, SD or retention money may be refunded at the earliest, if the contractor is not liable to pay any money to IREL under any other contract.

5.0 **FORFEITURE OF SD & RETENTION MONEY**

The SD & retention money shall stand forfeited in favour of IREL, without any further notice to the contractor in the following circumstances:

- (i) In case of any failure whatsoever on the part of the Supplier/ Contractor at any time during performance of his part of the contract including the extended periods of contract, where notice is given and time for rectification allowed.
- (ii) If the Supplier/ contractor indulges at any time in any subletting/ subcontracting of any portion of the work without approval of IREL.

PRE-QUALIFICATION CRITERIA (PQC)

Tenderer should meet the following minimum pre-qualification criteria:

1. TECHNICAL COMPETENCE

- 1.1 The Bidder should be OEM of HP Ge detector based gamma spectrometer or their authorized dealer and should have supplied similar item to any Central/State Govt. Organization/PSU/Public Listed Company/Reputed Private company in last 7 years ending on 31st March 2024. Bidder shall provide documentary evidence of copies of relevant supply documents/ purchase order along with copy of satisfactory installation report/proof of completion to whom the equipment was supplied for the value as follows.
- (i) Execution of at least one order for "similar work for a value of not less than Rs.155.19 lakh."

 OR
- (ii) Execution of two orders for "similar works for a value each not less than Rs.96.99 lakh."
- (iii) Execution of three orders for "similar works for a value each not less than Rs.77.59 lakh."

Experience in Similar item/work means Supply, Installation, Commissioning of HP Ge detector based gamma spectrometer with accessories

- 1.2 Bids shall be submitted by OEM or their Authorized Agency/Authorized Seller. (OEM means OEM of HP Ge detector based gamma spectrometer). Valid authorization certificate from OEM of HP Ge detector based gamma spectrometer, incase Authorized agency/ Authorized seller is to be submitted along with the technical bid.
- 1.3 Party has to furnish an undertaking that they will be able to carry-out AMC, supply all the required spare of all parts of HP Ge detector based gamma spectrometer for a minimum period of 10 years after completion of comprehensive warranty on payment / separate order basis.
- 1.4 The bidder has to confirm no end user certificate is required for the supply of HP Ge detector based gamma spectrometer. IREL (India) Limited will not issue the same.

2 FINANCIAL SOUNDNESS

The bidder shall have average annual financial turn over during the last 3 financial years ending on 31st March 2024 should be at least Rs.58.19 lakh to be supported by audited balance sheet & profit and loss account/ turnover certificate from audit firms/ CA for last three financial years.

Note:-

Documents against the Pre-qualification criteria as above complete in all respect must be uploaded. Failure to compliance of above shall make the bid incomplete and as such it shall be rejected. In the absence of receipt of Pre-qualification criteria, the offer will not be considered under any circumstances

Undertaking for AMC

We hereby confirm that we will be able to carry out AMC, supply the required spares for all items (all parts of HP Ge detector based gamma spectrometer, UPS, UPS Batteries, computer, printer, monitor for the computer etc) for a period of minimum 10 years after warranty period of Instrument on payment basis as per requirement.

Signature
Name
Name of the Company and Address (with Seal)

PART - I

TECHNO-COMMERCIAL BID

REQUIRED DOCUMENTS TO BE SUBMITTED ALONGWITH TECHNO-COMMERCIAL BID.

Sl. No.	Description	Up-loaded
		Yes/No
1	PROOF OF SUBMISSION OF EMD / VALID EXEMPTION	
	CERTIFICATE (If Applicable).	
2	DOCUMENTS AS PER PRE-QUALIFICATION CRITERIA.	
3	DULY FILLED, SIGNED AND SEAL AFFIXED IREL TENDER	
	DOCUMENTS (Techno-Commercial Bid) (Page 9 - 34).	
4	COPY OF GST & PAN CERTIFICATES.	
5	ANY OTHER DOCUMENTS AS PER TENDER CONDITIONS.	

SPECIAL CONDITIONS OF CONTRACT (SCOC)

1.0 SCOPE OF WORK

The scope of the work includes the following.

- 1.1 Supply, installation, testing, commissioning and operational training of HP Ge detector based gamma spectrometer as per the technical specification attached which includes required accessories with minimum 2 year comprehensive warranty.
- 1.2 Annual maintenance contract (AMC) for 3 years after the completion of comprehensive warranty period. The scope of AMC is described in clause no.10.

2.0 SCOPE OF IREL (India) Limited

- 2.1 IREL will provide the following materials / utilities free of cost
 - (a) Electricity will be made available at a single point from where the contractor shall make their own arrangements for using the same, following the safety regulations practiced in IREL.
 - (b) Power connections
 - (c) High pure water
 - (d) Bench/platform for installing the equipment

3.0 **CONTRACTOR'S SCOPE**

All materials required for the completion of Supply to commissioning of and training of HP Ge detector based gamma spectrometer is under the Contractor scope.

4.0 **PERIOD OF CONTRACT**

- 4.0 Supply, installation, testing, commissioning and training has to be completed within 270 days from the date of the order.
- 4.1 Permitted work timings inside factory premises is from 0800 hrs. to 1600 hrs. on all working days (excluding Sundays & holidays applicable to IREL). Any request for extension of working time will be granted only in case of exigencies under the sole discretion of EIC.
- 4.2 Any delay on the part of IREL in handing over the site or stoppage of work-in-progress due to reasons attributable to IREL will be noted separately and the date of completion will be extended accordingly.

5.0 **PAYMENT TERMS**

- 5.1 60% of total order value for the supply part less statutory deductions shall be released within 30 days of the receipt and acceptance of HP-Ge detector based gamma spectrometer including all accessories at IREL, Udyogamandal site and acceptance by O.I.C.
- 5.2 35% of total order value for the supply part less statutory deductions shall be released on successful completion of Installation, testing, commissioning & training duly certified by the OIC.
- 5.3 Balance 5% payment for the supply part along with SD amount after adjusting the SD amount for AMC value shall be released after comprehensive warranty period of 24 months from the date of completion of commissioning and training or against submission of BG for the equivalent amount and period. BG will be refunded after completion of 2 year comprehensive warranty period and on submission of SD amount for the AMC.
- 5.4 No advance payment will be made under any circumstances.

- 5.5 The Contractor's request(s) for payment shall be made to IREL in writing, accompanied by invoice in triplicate, describing, as appropriate, the equipment delivered and Services performed, inspection report submitted and fulfilment of other obligations stipulated in the Contract.
- 5.6 The Contractor has to submit Warrantee certificate for the complete system with accessories while submitting the final claim.
- 5.7 Payment shall be made after statutory deductions by NEFT/RTGS. All bills/invoices are to be certified by EIC. Party has to submit commercially clear invoice/ bills to avail Input Tax credit on the materials supplied.

MODE OF PAYMENT:

The payment shall be released after deducting the following:

- (a) Compensation recoverable if any.
- (b) LD recoverable, if any.
- (c) Recoveries on account of contractee's facilities and services, if any.
- (d) Deduction, if any, towards re-coupment of any amounts forfeited, adjusted, deducted out of security deposit.
- (e) All Statutory deduction, if any, as per rules.
- (f) Deduction as per IT Act, 1961, if any.
- (g) Any other deduction related to this contract.

6.0 **WARRANTY**

- 6.1 All the items (all parts of HP-Ge detector based gamma spectrometer, UPS, UPS Batteries, computer, printer, monitor for the computer etc.) supplied shall be under comprehensive warranty against any defective material, malfunctioning or poor workmanship for a period of two years from the date of installation, commissioning, testing of the instrument as detailed in Technical specification. The successful bidder will be responsible for the comprehensive maintenance of the supplied machine and its related accessories free of charge during the warranty period. During warranty period any software wrt operating system or its licence period renewal & application package and hardware part required to run the application package for carrying out the analysis shall be provided by the supplier at no additional cost.
- 6.2 The bidder should give undertaking to carry out AMC, supply the required spares for all items (all parts of HP-Ge detector based gamma spectrometer, UPS, UPS Batteries, computer, printer, monitor for the computer etc) for a period of minimum 10 years from the date of completion of two years comprehensive warranty period of Instrument on payment basis as per requirement.
- 6.3 During two years comprehensive warranty period transportation, boarding and Lodging charges of the service Engineers will be borne by the bidder.

7.0 **PENALTY**

In case of any complaint in the working of the Instrument during warranty period, the firm shall attend the same within 48 hours from the time of complaint call/ email by IREL and the instrument must be repaired within 10 days. Further, in case of any failure to do so, penalty @ Rs 2000/- per day will be applicable.

8.0 SAFETY, SUPERVISION & HOUSE KEEPING

- 8.1 The Supplier is fully responsible for carrying out the work in safe manner. Successful bidder has to carry out the work with utmost care and as per the rules laid down by IREL safety department.
- 8.2 The work must be carried out under close supervision and without causing any damage / hindrance to the existing structure or equipment's and other activities in the surrounding areas.

- 8.3 Special Work Permit/ Safety Permit shall be obtained for working at heights and other hazardous areas as specified by Safety department prior to commencement of work on each day.
- 8.4 The successful bidder must provide Personal Protective Equipment's like safety shoes, helmet, gloves, goggles etc., for their workmen suitable for the working environment and also as per the advice of Safety-in-charge / Engineer-in-charge.
- 8.5 Electrical extension boards are to be fitted with safety devices like ELCB, MCB/Fuse units etc. Each power tool shall be supplied through individual plug sockets. Electrical power shall be tapped through ISI marked pin plug top. Flexible cable used shall be of FRLS type with adequate current carrying capacity for the loads to be collected. Poor insulated cables and too many joints on the cable shall be avoided.

9.0 **SITE INSPECTION**

- 9.1 The bidders are advised to inspect the site before submitting their quote. The bidders are also advised to study the labour supply/availability, trade union practices and labour rates prevailing at Udyogamandal.
- 9.2 Once the tenderer submits his bid, the following are fully understood by both the parties (i.e. IREL and the Tenderer) binding on them:
- 9.2.1 Before the tenderer submits his offer, the tenderer is deemed to have carefully examined the Technical Specification, General and Special Conditions of Contract and other details relating to tender requirement and fully acquaint himself as to all conditions and matters which may in anyway affect the work or the cost thereof. The bidder shall be deemed to have obtained on his own and independently all necessary information for the purpose of preparing the bid and his bid as accepted shall be deemed to have taken into account all contingencies as may arise due to such information or lack of the same.
- 9.2.2 IREL shall not entertain any request for clarifications from the bidders regarding such local conditions. Any neglect or omission or failure on the part of the bidder in obtaining necessary and reliable information as stated above or on any other matter affecting the bidder shall not relieve him from any risks or liabilities or the entire responsibility for completion of the work in accordance with the tender document.
- 9.2.3 The bidder shall be deemed to have acquainted himself with the Indian Income Tax Act, 1961, Indian Companies Act, 1956, Indian Customs Act, 1962, Indian Electricity Act 2003, Factories Act 1948, Indian Mines Act 1952, Pollution Control Regulation and other related Acts & Laws prevalent in India and as amended from time to time.

10.0 ANNUAL MAINTENANCE CONTRACT (AMC) TERMS AND CONDITIONS

- 10.1 The scope includes general preventive maintenance as well as breakdown calls, Hardware and Software and any other problems related to the instrument and accessories supplied along with the HP Ge, UPS and its batteries. During AMC period any software wrt operating system or its licence period renewal & application package and hardware part required to run the application package for carrying out the analysis shall be provided by the supplier at no additional cost.
- 10.2 Other accessories required to run the instrument such as Computer, Printer, Gas, Stabiliser for computer are not covered under this contract.
- 10.3 AMC will be for 3 years after the completion of comprehensive warranty period of 2 years. Separate charges applicable for 1st, 2nd, 3rd years.
- 10.4 The quoted AMC charges per Annum will be inclusive of transport, boarding and Lodging of the service Engineers.
- 10.5 No separate order will be issued for AMC.
- 10.6 The number of preventive maintenance visits shall be minimum four (4) per annum.
- 10.7 In case of any breakdown party should take up the work within 48 hours of intimation. There is no restriction on breakdown calls during AMC period of 3 years. Any breakdown call charges will be borne by the bidder.
- 10.8 The supplier has to inform critical spares to be maintained during AMC period. Spares will be provided by IREL during AMC period.
- 10.9 AMC amount will be released as per AMC terms and conditions and payment shall be released within 30 days on receipt of invoice after deducting statutory levies (if any) after completion of every AMC visit in respective years or pro-rata basis and duly recommended by EIC.

10.10 PRE-BID MEETING

The Pre-bid meting shall be conducted on 04/04/2025 as online/offline. Online details will be send to the party based on their response through mail to IREL.

10.0 OFFICER-IN-CHARGE (OIC)

Smt. Neetha Rajesh, M (QC&R) will be the Officer-in-Charge for this work.

11.0 NO END USER CERTIFICATE SHALL BE ISSUED BY IREL (India) Limited

13.0 **BILL OF QUANTITIES (BOQ)**

Sl. No	Item Description	Unit	Quantity
	Supply, installation, commissioning, testing & training of		
1	HPGe detector based gamma spectrometer and accessories		
1	with minimum two years comprehensive warranty as	No.	1
	detailed in the specification in Annexure-IIA	1 (0.	
2	AMC amount for 1st year	No.	1
3	AMC amount for 2 nd year	No.	1
4	AMC amount for 3 rd year	No.	1

14.0 <u>DETAILS TO BE FURNISHED ALONG WITH THE OFFER</u>

Sl. No.	Item Description	GST in %	HSN Code
1	HP Ge detector based gamma spectrometer		
2	AMC for 3 years after 2 years warranty period		

Price Bid Evaluation basis for tender:

- L1 price evaluation of the technically qualified bids as per pre-qualification criteria shall be done based on the total landed price in INR at IREL, RED, Udyogamandal, Kerala.
- For arriving L1 price towards placement of order through GEM, total cost including AMC charges will be considered.

Technical Specifications for HPGe based Gamma Ray Spectrometer

S. No	Description	Value	Compliance
1.00	Detector		-
1.01	P-type coaxial vertically vacuum dipstick mount (HP_Ge) with pre-amplifier housed in a single	Ç ,	
1.02	End Cap diameter	76mm (3.2 inch) dia	
1.03	Input Window material	Carbon Composite	
1.04	Energy Range	3 keV to 10 MeV	
1.05	Relative Efficiency	≥ 40 %	
1.06	Energy Resolution as FWHM at 122 keV	≤ 0.875 keV	
1.07	Energy Resolution as FWHM at 1.33 MeV	≤1.80 keV	
1.08	Peak shape (FWTM/FWHM)	≤ 1.90 keV	
1.09	Peak to Compton Ratio	≥62:1	
1.10	HV Inhibit (Warm-Up/HV shutdown)	Must	
1.11	Should provide Detector test/performance cert		
	Cables Set for Power, HV, Output Signal, Pre-		
1.12	Amplifier and LN2 Sensor	1 Set	
1.13	Radio Active Source KIT for Detector Calibration	GDB-1: Set of 5 (Ba-133; Cs-137; Co-57; Co-60;Na-22) Disk Type Gamma Reference & Dummy Source, Source Strength is nearly 3 Micro Cu source and Eu-152 source with 0.1 Micro Cu Strength	
2.00	Electrical Cooling Systems (Cp-5) Specification	ons (Integrated)	1 Number
2.01	Power Controller:	 Dimensions - 280 x 88 x 315 mm (11 x 3.5 x 12.4 in.) Weight - 5.3 kg (11.6 lb). Power Controller Cable Length - 3 m (10 ft). 	
2.02	Power Requirements:	 100-240 V ac, 50-60 Hz (auto ranging power supply). < 100 W nominal < 200 W max. (during cool down) Fuse - 2x T3.15 A 250 V 	
2.03	Connectors:	 2X BNC-F - HV-inhibit. RS-232 - Remote control and status read-out (USB/RS-232 adapter provided). DB15-F - Relay output - cooler status and warm/cold indication. 	
2.04	MTTF (Mean Time to Failure)	> 3,000,000 hoursRemote read-outPulse-tube technology	
2.05	Operating Temperature:	+5 to +40 °C (41 to 104 °F) on standard models and configurations	
2.06	Cold Head: (Excluding detector chamber)	 Dimensions - 145 x 287 x 313 mm (5.7 x 11.3 x 12.3 in.) (W x H x D). Weight - 17 kg (37.5 lb) approx. 	
3.00	Pre-Amplifier		
3.01	Intelligent Pre-Amplifier built in Detector Caps		
3.02	Pre-Amplifier parameters	Can be studied using software	
3.03	Pre-amplifier software should provide the	Detector Leakage Current	

	parameters	Charge Loop DC level	
	r	Out stage DC level	
		FET Backgate Voltage	
		FET Drain Ref Voltage	
		HV Inhibit status	
		High Count Rate status	
İ		PRTD1 and PRTD2 Temperature	
İ		Voltage supply	
İ		Can adjust Pole Zero	
İ		Can adjust PRTD set points	
3.04	Pre-amplifier should be connected to PC using	, , , , , , , , , , , , , , , , , , , ,	
4.00	Digital Signal Analyzer	COD to study its parameters	1 Number
1,00	Integrated Desktop Multi Channel Analyzer ba	sed Digital Signal Analyzer Operates through	1114111201
4.01	a software and compatible with the detector.	isea Digital Signal Many Zer. Operates unough	
1.01	It comprises:		
4.02	16 K (Minimum) Channel conversion gain/spe	ctrum memory	
4.03	Communication type & Speed	USB with 12 Mbits/s	
4.04	Modes of Operation	MCS/PHA mode	
1,01	modes of operation	Live or True Time, counts in single channel,	
		Counts in ROI, Counts in multiple ROIs	
4.05	PHA Mode	256, 512, 1024, 2048, 4096, 8192 or 16 384	
1.00	1111 Wode	channels. Support for two memory groups of	
İ		equal size	
		Sweeps, count greater or equal to preset	
		Counts, count greater or equal to preset ROI	
		Counts	
4.06	MCS Mode	256, 512, 1024, 2048, 4096, 8192 or 16 384	
İ		channels. Support for two memory groups of	
		equal size	
4.07	Integrated desktop MCA based on Digital Sign	1	
4.08	Count Rate and Temperature stability	, , , , , , , , , , , , , , , , , , ,	
4.09	Rise time	0.2 to 38 μs.	
4.10	Flat Top	0 to 3 μs	
4.11	Coarse Gain	x2 .0 – x430.5 in 19% increments	
4.12	Fine Gain	$x0.8$ to $x1.2$ in $\sim 0.004\%$ increments	
		≤±0.025% of full scale over the top 99% of	
4.13	Integral Non-Linearity	selected range.	
		≤±1% over the top 99% of the range including	
4.14	Differential Non-Linearity	the effects from integral non-linearity	
4.15	Gain Drift	≤35 ppm/°C after 15 minutes of operation.	
4.16	Zero Drift	<3 ppm/°C after 15 minutes of operation	
		Better than 500 ns with NORMAL Fast	
4.17	Pulse pair resolution	Discriminator mode selected.	
4.18	PUR Guard	1.1 to 2.5 in 0.1 increments	
4.19	Indicators	Full set of Front panel LED indicators	
4.20	Auto pole zero correction facility	Yes	
4.21	Input polarity	Software Selectable	
4.22	Live Time Correction facility	Yes	
4.23	Digital Stabilization Capability	Yes	
	Base line restorer	AUTO, HARD, MEDIUM, SOFT	
	L Dase line resioner		
4.24			
	Digital fast discriminator	Yes	
4.24			

Ranges: 1200 to 1500 V dc or x1500 to 45000 V dc Ranges: 10 m V P-P Temperature Coefficient; 450 ppm/°C. Subility: 2018 /h, 0,023 /h b. Accuracy: 545% Load Regulation: 1% Setting Resolution:12-bit (1/4096) 4.28 Operates directly or via power adapter from 200 to 240 V AC 50 Hz power supply 4.29 Overload protected Detector Preamp Power Supply. 4.20 Overload protected Detector Preamp Power Supply. 4.21 V ±5% at 50 mA max Supply. 4.22 Overload protected Detector Preamp Power Supply. 4.23 All Metal Enclosure Supply. 4.24 V ±5% at 100 mA max The Integrated Desktop Multi Channel Analyzer communicates with computer USB port Port of the Power Supply. 4.22 Operating Temperature All Metal Enclosure All Metal Enclosure All Metal Enclosure All Metal Enclosure All Metal Enclosure All Metal Enclosure All Metal Enclosure All Operating Humidity S5% non-condensing as per FN-61010 5.00 Gamma Spectroscopy Analysis Software, comprises of following Proloaded with Insteat version of instrument software compatible and operates/control all the functions of: 1. Digital Signal Analyzer, 2. Pre amplifier and 3. HP Go Detector & Cryostat 4. Control High Voltage Power Supply 5.02 Peak Search & Hibrary Directed Peak Search 5.03 Peak Area calculations All Multiple multiple spectrums is a single window Multiple multiple spectrums is a single window Compare two spectrums in a single window Read Reak-ground Subtraction Standard Background Subtraction feature Automatic Interference correction 5.00 Each Search Subtraction Standard Background Subtraction feature Automatic Interference correction 5.01 Peak Smoothing and Stripping feature 5.02 Automatic Interference correction 5.03 Can able to view multiple spectrums good to the single spectrums of the Search 5.04 Automatic Interference correction 5.05 Can be the view of the single spectrums counting time should not get added while merging 5.08 Peak Smoothing and Stripping feature 5.09 Ref Background subtraction Standard Background Subtraction feature 5.00 Report generation in printable form		1	Doloviter Coloratable by Coltmans	
V de Ripple: 31 m V P-P Temperature Coefficient ±50 ppm/°C.			Polarity: Selectable by Software	
Temperature Coefficient: ±50 ppm/°C. Stability: 0.01%/h, 0.02%/8 h. Accuracy: ≤ ±5% Load Regulation: 13%			e e e e e e e e e e e e e e e e e e e	
Temperature Coefficient: ±50 ppm/°C. Stability: 0.01%/h, 0.02%/8 h. Accuracy: ≤ ±5% Load Regulation: 13%			Ripple: ≤ 10 mV P-P	
Stability: 0.018 /h. 0.028 /8 h. Accuracy: ≤ £5% Load Regulation: 1%			11	
Accuracy: \$45%				
Load Regulation: 1%				
Setting Resolution:12-bit (1/4096)				
4.28 Operates directly or via power adapter from 200 to 240 V AC 50 Hz power supply				
A.29 Overload protected Detector Preamp Power ±24 V ±5% at 50 mA max	4.28	Operates directly or via power adapter from 20		
4.30 Communication Analyzer communicates with computer USB port 4.31 Physical All Metal Enclosure 4.32 Operating Temperature 4.33 Operating Temperature 4.34 Operating Humidity 5.00 Gamma Spectroscopy Analysis Software, comprises of following Preloaded with latest version of instrument software compatible and operates/control all the functions of: 1. Digital Signal Analyzer, 2. Pre amplifier and 3. HP Ge Detector & Cryostat 4. Control High Voltage Power Supply 5.02 Peak Search & Library Directed Peak Search 5.03 Peak Area calculations 5.04 Multiple methods for Energy & Efficiency Calibration 5.05 Can able to view multiple spectrums side by side 5.06 Compare two spectrums in a single window 5.07 Can merge multiple spectrums; counting time should not get added while merging 5.08 Peak Smoothing and Stripping feature 5.09 Peak Smoothing and Stripping feature 5.09 Ref. Background subtraction, Standard Background Subtraction feature 5.10 Automatic Interference correction 5.11 Parent/ daughter decay correction 5.12 Gain Evaluator & Auto Gain Adjustment 5.13 Nuclide Activity Calculation 5.14 Minimum Detectable Activity Calculation 5.15 Peak Weighted mean Nuclidie activity calculation 5.16 Auto Nuclide identification 5.17 Report generation in printable format 5.18 Hard Copy and complete documents of the Licensed Software 5.19 Quality assurance feature to monitor and record parameters such as peak centroid, FWHM, Nuclide activities, MDA and other important results. 5.20 Ability to review, adjust, or add/delete peaks in specific regions/ROI 5.21 Ability to review, adjust, or add/delete peaks in specific regions/ROI 5.24 Should be organized into multiple sectioned based pages 5.20 Should provide information such as basic sample information; a spectral image with count				
Communication	4.29	_		
Analyzer communicates with computer USB port		Supply.		
Physical All Metal Enclosure	4.20	Communication		
4.32 Operating Temperature -20 to 50 °C	4.30	Communication	_	
4.32 Operating Temperature .20 to 50 °C	4.01	Dl 1	<u> </u>	
4.33 Operating Humidity 85% non-condensing as per EN-61010		ž		
February February				
Preloaded with latest version of instrument software compatible and operates/control all the functions of: 1. Digital Signal Analyzer, 2. Pre amplifier and 3. HP. Ge Detector & Cryostat 4. Control High Voltage Power Supply 5.02 Peak Search & Library Directed Peak Search 5.03 Peak Area calculations 5.04 Multiple methods for Energy & Efficiency Calibration 5.05 Can able to view multiple spectrums side by side 5.06 Compare two spectrums in a single window 5.07 Can merge multiple spectrums; counting time should not get added while merging 5.08 Peak Smoothing and Stripping feature 5.09 Ref. Background subtraction, Standard Background Subtraction feature 5.10 Automatic Interference correction 5.11 Parent/daughter decay correction 5.12 Gain Evaluator & Auto Gain Adjustment 5.13 Nuclide Activity Calculation 5.14 Minimum Detectable Activity Calculations with Curie MDA, KTA MDA, ISO11929 MDA 5.15 Peak Weighted mean Nuclide activity calculation 5.16 Auto Nuclide identification 5.17 Report generation in printable format 5.18 Hard Copy and complete documents of the Licensed Software 5.20 Quantification in units μCi/Bq Quality assurance feature to monitor and record parameters such as peak centroid, FWHM, Nuclide activities, MDA and other important results. The QA data stored and the results obtained should be displayed in a chart type or as a report template. 5.21 Ability to review, adjust, or add/delete peaks in specific regions/ROI 5.22 Python scripting integration for customizable results. Results review and reporting in software 5.23 Should be organized into multiple sectioned based pages 5.34 Should provide information such as basic sample information; a spectral image with count		1 0	Ŭ I	4 37 1
functions of: 1. Digital Signal Analyzer, 2. Pre amplifier and 3. HP_Ge Detector & Cryostat 4. Control High Voltage Power Supply 5.02 Peak Search & Library Directed Peak Search Peak Area calculations 5.04 Multiple methods for Energy & Efficiency Calibration 5.05 Can able to view multiple spectrums side by side 5.06 Compare two spectrums in a single window 5.07 Can merge multiple spectrums; counting time should not get added while merging 5.08 Peak Smoothing and Stripping feature 5.09 Ref. Background subtraction, Standard Background Subtraction feature 5.10 Automatic Interference correction 5.11 Parent/daughter decay correction 5.12 Gain Evaluator & Auto Gain Adjustment 5.13 Nuclide Activity Calculation 5.14 Minimum Detectable Activity Calculation 5.15 Peak Weighted mean Nuclide activity calculation 5.16 Auto Nuclide identification 5.17 Report generation in printable format 5.18 Hard Copy and complete documents of the Licensed Software 5.19 Quantification in units µCi/Bq Quality assurance feature to monitor and record parameters such as peak centroid, FWHM, Nuclide activities, MDA and other important results. The QA data stored and the results obtained should be displayed in a chart type or as a report template. 5.20 Python scripting integration for customizable results. Results review and reporting in software 5.21 Ability to review, adjust, or add/delete peaks in specific regions/ROI 5.22 Should be organized into multiple sectioned based pages 5.24 Should be organized into multiple sectioned based pages 5.24 Should be organized into multiple sectioned based pages 5.24 Should be organized into multiple sectioned based pages 5.24 Should be information such as basic sample information; a spectral image with count	5.00			1 Number
1. Digital Signal Analyzer, 2. Pre amplifier and 3. HP_Ce Detector & Cryostat 4. Control High Voltage Power Supply 5.02 Peak Search & Library Directed Peak Search 5.03 Peak Area calculations 5.04 Multiple methods for Energy & Efficiency Calibration 5.05 Can able to view multiple spectrums side by side 5.06 Compare two spectrums in a single window 5.07 Can merge multiple spectrums; counting time should not get added while merging 5.08 Peak Smoothing and Stripping feature 5.09 Ref. Background subtraction, Standard Background Subtraction feature 5.10 Automatic Interference correction 5.11 Parent/ daughter decay correction 5.12 Gain Evaluator & Auto Gain Adjustment 5.13 Nuclide Activity Calculation 5.14 Minimum Detectable Activity Calculations with Curie MDA, KTA MDA, ISO11929 MDA 5.15 Peak Weighted mean Nuclide activity calculation 5.16 Auto Nuclide identification 5.17 Report generation in printable format 5.18 Hard Copy and complete documents of the Licensed Software 5.19 Quantification in units µCi/Bq Quality assurance feature to monitor and record parameters such as peak centroid, FWHM, Nuclide activities, MDA and other important results. The QA data stored and the results obtained should be displayed in a chart type or as a report template. 5.20 Python scripting integration for customizable results. Results review and reporting in software - Subuld be organized into multiple sectioned based pages 5.24 Should be organized into multiple sectioned based pages 5.24 Should be organized into multiple sectioned based pages 5.24 Should be organized into multiple sectioned based pages 5.24 Should be organized into multiple sectioned based pages 5.24 Should be organized into multiple sectioned based pages 5.25 Should be organized into multiple sectioned based pages 5.26 Should be organized into multiple sectioned based pages 5.27 Should be organized into multiple sectioned based pages 5.28 Should be organized into multiple sectioned based pages			tware compatible and operates/control all the	
2. Pre amplifier and 3. HP_Ge Detector & Cryostat 4. Control High Voltage Power Supply 5.02 Peak Search & Library Directed Peak Search 5.03 Peak Area calculations 5.04 Multiple methods for Energy & Efficiency Calibration 5.05 Can able to view multiple spectrums side by side 5.06 Compare two spectrums in a single window 5.07 Can merge multiple spectrums; counting time should not get added while merging 5.08 Peak Smoothing and Stripping feature 5.09 Ref. Background subtraction, Standard Background Subtraction feature 5.10 Automatic Interference correction 5.11 Parent/ daughter decay correction 5.12 Gain Evaluator & Auto Gain Adjustment 5.13 Nuclide Activity Calculation 5.14 Minimum Detectable Activity Calculations with Curie MDA, KTA MDA, ISO11929 MDA 5.15 Peak Weighted mean Nuclide activity calculation 5.16 Auto Nuclide identification 5.17 Report generation in printable format 5.18 Hard Copy and complete documents of the Licensed Software 5.19 Quantification in units μCi/Bq Quality assurance feature to monitor and record parameters such as peak centroid, FWHM, Nuclide activities, MDA and other important results. The QA data stored and the results obtained should be displayed in a chart type or as a report template. 5.20 Ability to review, adjust, or add/delete peaks in specific regions/ROI 5.21 Ability to review, adjust, or add/delete peaks in specific regions/ROI 5.22 Python scripting integration for customizable results. Results review and reporting in software 5.23 Should be organized into multiple sectioned based pages 5.24 Should provide information such as basic sample information; a spectral image with count				
3. HP_Ge Detector & Cryostat 4. Control High Voltage Power Supply 5.02 Peak Search & Library Directed Peak Search 5.03 Peak Area calculations 5.04 Multiple methods for Energy & Efficiency Calibration 5.05 Can able to view multiple spectrums side by side 5.06 Compare two spectrums in a single window 5.07 Can merge multiple spectrums; counting time should not get added while merging 5.08 Peak Smoothing and Stripping feature 5.09 Ref. Background subtraction, Standard Background Subtraction feature 5.10 Automatic Interference correction 5.11 Parent/daughter decay correction 5.12 Gain Evaluator & Auto Gain Adjustment 5.13 Nuclide Activity Calculation 5.14 Minimum Detectable Activity Calculations with Curie MDA, KTA MDA, ISO11929 MDA 5.15 Peak Weighted mean Nuclide activity calculation 5.16 Auto Nuclide identification 5.17 Report generation in printable format 5.18 Hard Copy and complete documents of the Licensed Software 5.19 Quantification in units μCi/Bq Quality assurance feature to monitor and record parameters such as peak centroid, FWHM, Nuclide activities, MDA and other important results. The QA data stored and the results obtained should be displayed in a chart type or as a report template. 5.21 Ability to review, adjust, or add/delete peaks in specific regions/ROI 5.24 Should be organized into multiple sectioned based pages 5.24 Should be organized into multiple sectioned based pages 5.24 Should be organized into multiple sectioned based pages 5.24 Should provide information such as basic sample information; a spectral image with count	5.01			
4. Control High Voltage Power Supply 5.02 Peak Search & Library Directed Peak Search 5.03 Peak Area calculations 5.04 Multiple methods for Energy & Efficiency Calibration 5.05 Can able to view multiple spectrums side by side 5.06 Compare two spectrums in a single window 5.07 Can merge multiple spectrums; counting time should not get added while merging 5.08 Peak Smoothing and Stripping feature 5.09 Ref. Background subtraction, Standard Background Subtraction feature 5.10 Automatic Interference correction 5.11 Parent/daughter decay correction 5.12 Gain Evaluator & Auto Gain Adjustment 5.13 Nuclide Activity Calculation 5.14 Minimum Detectable Activity Calculations with Curie MDA, KTA MDA, ISO11929 MDA 5.15 Peak Weighted mean Nuclide activity calculation 5.16 Auto Nuclide identification 5.17 Report generation in printable format 5.18 Hard Copy and complete documents of the Licensed Software 5.19 Quantification in units μCi/Bq Quality assurance feature to monitor and record parameters such as peak centroid, FWHM, Nuclide activities, MDA and other important results. The QA data stored and the results obtained should be displayed in a chart type or as a report template. 5.20 Python scripting integration for customizable results. - Results review and reporting in software - Should be organized into multiple sectioned based pages 5.24 Should be organized into multiple sectioned based pages 5.24 Should be organized into multiple sectioned based pages 5.24 Should provide information such as basic sample information; a spectral image with count				
5.02 Peak Search & Library Directed Peak Search 5.03 Peak Area calculations 5.04 Multiple methods for Energy & Efficiency Calibration 5.05 Can able to view multiple spectrums side by side 5.06 Compare two spectrums in a single window 5.07 Can merge multiple spectrums; counting time should not get added while merging 5.08 Peak Smoothing and Stripping feature 5.09 Ref. Background subtraction, Standard Background Subtraction feature 5.10 Automatic Interference correction 5.11 Parent/daughter decay correction 5.12 Gain Evaluator & Auto Gain Adjustment 5.13 Nuclide Activity Calculation 5.14 Minimum Detectable Activity Calculations with Curie MDA, KTA MDA, ISO11929 MDA 5.15 Peak Weighted mean Nuclide activity calculation 5.16 Auto Nuclide identification 5.17 Report generation in printable format 5.18 Hard Copy and complete documents of the Licensed Software 5.19 Quantification in units µCi/Bq Quality assurance feature to monitor and record parameters such as peak centroid, FWHM, Nuclide activities, MDA and other important results. 5.20 The QA data stored and the results obtained should be displayed in a chart type or as a report template. 5.21 Ability to review, adjust, or add/delete peaks in specific regions/ROI 5.22 Python scripting integration for customizable results. 6.23 Should be organized into multiple sectioned based pages 5.24 Should provide information such as basic sample information; a spectral image with count				
 5.03 Peak Area calculations 5.04 Multiple methods for Energy & Efficiency Calibration 5.05 Can able to view multiple spectrums side by side 5.06 Compare two spectrums in a single window 5.07 Can merge multiple spectrums; counting time should not get added while merging 5.08 Peak Smoothing and Stripping feature 5.09 Ref. Background subtraction, Standard Background Subtraction feature 5.10 Automatic Interference correction 5.11 Parent/daughter decay correction 5.12 Gain Evaluator & Auto Gain Adjustment 5.13 Nuclide Activity Calculation 5.14 Minimum Detectable Activity Calculations with Curie MDA, KTA MDA, ISO11929 MDA 5.15 Peak Weighted mean Nuclide activity calculation 5.16 Auto Nuclide identification 5.17 Report generation in printable format 5.18 Hard Copy and complete documents of the Licensed Software 5.19 Quantification in units μCi/Bq Quantification in units μCi/Bq Quality assurance feature to monitor and record parameters such as peak centroid, FWHM, Nuclide activities, MDA and other important results. 5.20 The QA data stored and the results obtained should be displayed in a chart type or as a report template. 5.21 Ability to review, adjust, or add/delete peaks in specific regions/ROI 5.22 Python scripting integration for customizable results. 5.23 Should be organized into multiple sectioned based pages 5.24 Should provide information such as basic sample information; a spectral image with count 	5 00			
 5.04 Multiple methods for Energy & Efficiency Calibration 5.05 Can able to view multiple spectrums side by side 5.06 Compare two spectrums in a single window 5.07 Can merge multiple spectrums; counting time should not get added while merging 5.08 Peak Smoothing and Stripping feature 5.09 Ref. Background subtraction, Standard Background Subtraction feature 5.10 Automatic Interference correction 5.11 Parent/daughter decay correction 5.12 Gain Evaluator & Auto Gain Adjustment 5.13 Nuclide Activity Calculation 5.14 Minimum Detectable Activity Calculations with Curie MDA, KTA MDA, ISO11929 MDA 5.15 Peak Weighted mean Nuclide activity calculation 5.16 Auto Nuclide identification 5.17 Report generation in printable format 5.18 Hard Copy and complete documents of the Licensed Software 5.19 Quantification in units μCi/Bq Quantification in units μCi/Bq Quality assurance feature to monitor and record parameters such as peak centroid, FWHM, Nuclide activities, MDA and other important results. The QA data stored and the results obtained should be displayed in a chart type or as a report template. 5.21 Ability to review, adjust, or add/delete peaks in specific regions/ROI 5.22 Python scripting integration for customizable results. 5.23 Should be organized into multiple sectioned based pages 5.24 Should provide information such as basic sample information; a spectral image with count 		7		
5.05 Can able to view multiple spectrums side by side 5.06 Compare two spectrums in a single window 5.07 Can merge multiple spectrums; counting time should not get added while merging 5.08 Peak Smoothing and Stripping feature 5.09 Ref. Background subtraction, Standard Background Subtraction feature 5.10 Automatic Interference correction 5.11 Parent/daughter decay correction 5.12 Gain Evaluator & Auto Gain Adjustment 5.13 Nuclide Activity Calculation 5.14 Minimum Detectable Activity Calculations with Curie MDA, KTA MDA, ISO11929 MDA 5.15 Peak Weighted mean Nuclide activity calculation 5.16 Auto Nuclide identification 5.17 Report generation in printable format 5.18 Hard Copy and complete documents of the Licensed Software 5.19 Quantification in units µCi/Bq Quality assurance feature to monitor and record parameters such as peak centroid, FWHM, Nuclide activities, MDA and other important results. The QA data stored and the results obtained should be displayed in a chart type or as a report template. 5.21 Ability to review, adjust, or add/delete peaks in specific regions/ROI 5.22 Python scripting integration for customizable results. Results review and reporting in software 5.23 Should be organized into multiple sectioned based pages 5.24 Should provide information such as basic sample information; a spectral image with count			1	
5.06 Compare two spectrums in a single window 5.07 Can merge multiple spectrums; counting time should not get added while merging 5.08 Peak Smoothing and Stripping feature 5.09 Ref. Background subtraction, Standard Background Subtraction feature 5.10 Automatic Interference correction 5.11 Parent/daughter decay correction 5.12 Gain Evaluator & Auto Gain Adjustment 5.13 Nuclide Activity Calculation 5.14 Minimum Detectable Activity Calculations with Curie MDA, KTA MDA, ISO11929 MDA 5.15 Peak Weighted mean Nuclide activity calculation 5.16 Auto Nuclide identification 5.17 Report generation in printable format 5.18 Hard Copy and complete documents of the Licensed Software 5.19 Quantification in units μCi/Bq Quality assurance feature to monitor and record parameters such as peak centroid, FWHM, Nuclide activities, MDA and other important results. The QA data stored and the results obtained should be displayed in a chart type or as a report template. 5.21 Ability to review, adjust, or add/delete peaks in specific regions/ROI 5.22 Python scripting integration for customizable results. - Results review and reporting in software <		1 0, 1		
5.07 Can merge multiple spectrums; counting time should not get added while merging 5.08 Peak Smoothing and Stripping feature 5.09 Ref. Background subtraction, Standard Background Subtraction feature 5.10 Automatic Interference correction 5.11 Parent/daughter decay correction 5.12 Gain Evaluator & Auto Gain Adjustment 5.13 Nuclide Activity Calculation 5.14 Minimum Detectable Activity Calculations with Curie MDA, KTA MDA, ISO11929 MDA 5.15 Peak Weighted mean Nuclide activity calculation 5.16 Auto Nuclide identification 5.17 Report generation in printable format 5.18 Hard Copy and complete documents of the Licensed Software 5.19 Quantification in units μCi/Bq Quality assurance feature to monitor and record parameters such as peak centroid, FWHM, Nuclide activities, MDA and other important results. 5.20 The QA data stored and the results obtained should be displayed in a chart type or as a report template. 5.21 Ability to review, adjust, or add/delete peaks in specific regions/ROI 5.22 Python scripting integration for customizable results. - Results review and reporting in software 5.23 Should be organized into multip		1 1	de	
5.08 Peak Smoothing and Stripping feature 5.09 Ref. Background subtraction, Standard Background Subtraction feature 5.10 Automatic Interference correction 5.11 Parent/daughter decay correction 5.12 Gain Evaluator & Auto Gain Adjustment 5.13 Nuclide Activity Calculation 5.14 Minimum Detectable Activity Calculations with Curie MDA, KTA MDA, ISO11929 MDA 5.15 Peak Weighted mean Nuclide activity calculation 5.16 Auto Nuclide identification 5.17 Report generation in printable format 5.18 Hard Copy and complete documents of the Licensed Software 5.19 Quantification in units μCi/Bq Quality assurance feature to monitor and record parameters such as peak centroid, FWHM, Nuclide activities, MDA and other important results. The QA data stored and the results obtained should be displayed in a chart type or as a report template. 5.21 Ability to review, adjust, or add/delete peaks in specific regions/ROI 5.22 Python scripting integration for customizable results. - Results review and reporting in software 5.23 Should be organized into multiple sectioned based pages 5.24 Should provide information such as basic sample information; a spectral im		<u> </u>	1 11	
5.09 Ref. Background subtraction, Standard Background Subtraction feature 5.10 Automatic Interference correction 5.11 Parent/daughter decay correction 5.12 Gain Evaluator & Auto Gain Adjustment 5.13 Nuclide Activity Calculation 5.14 Minimum Detectable Activity Calculations with Curie MDA, KTA MDA, ISO11929 MDA 5.15 Peak Weighted mean Nuclide activity calculation 5.16 Auto Nuclide identification 5.17 Report generation in printable format 5.18 Hard Copy and complete documents of the Licensed Software 5.19 Quantification in units μCi/Bq Quality assurance feature to monitor and record parameters such as peak centroid, FWHM, Nuclide activities, MDA and other important results. The QA data stored and the results obtained should be displayed in a chart type or as a report template. 5.21 Ability to review, adjust, or add/delete peaks in specific regions/ROI 5.22 Python scripting integration for customizable results. - Results review and reporting in software 5.23 Should be organized into multiple sectioned based pages 5.24 Should provide information such as basic sample information; a spectral image with count				
5.10 Automatic Interference correction 5.11 Parent/daughter decay correction 5.12 Gain Evaluator & Auto Gain Adjustment 5.13 Nuclide Activity Calculation 5.14 Minimum Detectable Activity Calculations with Curie MDA, KTA MDA, ISO11929 MDA 5.15 Peak Weighted mean Nuclide activity calculation 5.16 Auto Nuclide identification 5.17 Report generation in printable format 5.18 Hard Copy and complete documents of the Licensed Software 5.19 Quantification in units μCi/Bq Quality assurance feature to monitor and record parameters such as peak centroid, FWHM, Nuclide activities, MDA and other important results. The QA data stored and the results obtained should be displayed in a chart type or as a report template. 5.21 Ability to review, adjust, or add/delete peaks in specific regions/ROI 5.22 Python scripting integration for customizable results. - Results review and reporting in software 5.23 Should be organized into multiple sectioned based pages 5.24 Should provide information such as basic sample information; a spectral image with count		0 11 0		
5.11 Parent/daughter decay correction 5.12 Gain Evaluator & Auto Gain Adjustment 5.13 Nuclide Activity Calculation 5.14 Minimum Detectable Activity Calculations with Curie MDA, KTA MDA, ISO11929 MDA 5.15 Peak Weighted mean Nuclide activity calculation 5.16 Auto Nuclide identification 5.17 Report generation in printable format 5.18 Hard Copy and complete documents of the Licensed Software 5.19 Quantification in units μCi/Bq Quality assurance feature to monitor and record parameters such as peak centroid, FWHM, Nuclide activities, MDA and other important results. The QA data stored and the results obtained should be displayed in a chart type or as a report template. 5.21 Ability to review, adjust, or add/delete peaks in specific regions/ROI 5.22 Python scripting integration for customizable results. - Results review and reporting in software 5.23 Should be organized into multiple sectioned based pages 5.24 Should provide information such as basic sample information; a spectral image with count			ound Subtraction feature	
5.12 Gain Evaluator & Auto Gain Adjustment 5.13 Nuclide Activity Calculation 5.14 Minimum Detectable Activity Calculations with Curie MDA, KTA MDA, ISO11929 MDA 5.15 Peak Weighted mean Nuclide activity calculation 5.16 Auto Nuclide identification 5.17 Report generation in printable format 5.18 Hard Copy and complete documents of the Licensed Software 5.19 Quantification in units μCi/Bq Quality assurance feature to monitor and record parameters such as peak centroid, FWHM, Nuclide activities, MDA and other important results. The QA data stored and the results obtained should be displayed in a chart type or as a report template. 5.21 Ability to review, adjust, or add/delete peaks in specific regions/ROI 5.22 Python scripting integration for customizable results. - Results review and reporting in software - 5.23 Should be organized into multiple sectioned based pages 5.24 Should provide information such as basic sample information; a spectral image with count				
 Nuclide Activity Calculation Minimum Detectable Activity Calculations with Curie MDA, KTA MDA, ISO11929 MDA Peak Weighted mean Nuclide activity calculation Auto Nuclide identification Report generation in printable format Hard Copy and complete documents of the Licensed Software Quantification in units μCi/Bq Quality assurance feature to monitor and record parameters such as peak centroid, FWHM, Nuclide activities, MDA and other important results. The QA data stored and the results obtained should be displayed in a chart type or as a report template. Ability to review, adjust, or add/delete peaks in specific regions/ROI Python scripting integration for customizable results. Results review and reporting in software Should be organized into multiple sectioned based pages Should provide information such as basic sample information; a spectral image with count 				
 5.14 Minimum Detectable Activity Calculations with Curie MDA, KTA MDA, ISO11929 MDA 5.15 Peak Weighted mean Nuclide activity calculation 5.16 Auto Nuclide identification 5.17 Report generation in printable format 5.18 Hard Copy and complete documents of the Licensed Software 5.19 Quantification in units μCi/Bq Quality assurance feature to monitor and record parameters such as peak centroid, FWHM, Nuclide activities, MDA and other important results. The QA data stored and the results obtained should be displayed in a chart type or as a report template. 5.21 Ability to review, adjust, or add/delete peaks in specific regions/ROI 5.22 Python scripting integration for customizable results. Results review and reporting in software 5.23 Should be organized into multiple sectioned based pages Should provide information such as basic sample information; a spectral image with count 		,		
 Feak Weighted mean Nuclide activity calculation Auto Nuclide identification Report generation in printable format Hard Copy and complete documents of the Licensed Software Quantification in units μCi/Bq Quality assurance feature to monitor and record parameters such as peak centroid, FWHM, Nuclide activities, MDA and other important results. The QA data stored and the results obtained should be displayed in a chart type or as a report template. Ability to review, adjust, or add/delete peaks in specific regions/ROI Python scripting integration for customizable results. Results review and reporting in software Should be organized into multiple sectioned based pages Should provide information such as basic sample information; a spectral image with count 		7		
 5.16 Auto Nuclide identification 5.17 Report generation in printable format 5.18 Hard Copy and complete documents of the Licensed Software 5.19 Quantification in units μCi/Bq Cuality assurance feature to monitor and record parameters such as peak centroid, FWHM, Nuclide activities, MDA and other important results. The QA data stored and the results obtained should be displayed in a chart type or as a report template. 5.21 Ability to review, adjust, or add/delete peaks in specific regions/ROI 5.22 Python scripting integration for customizable results. Results review and reporting in software 5.23 Should be organized into multiple sectioned based pages 5.24 Should provide information such as basic sample information; a spectral image with count 				
 5.17 Report generation in printable format 5.18 Hard Copy and complete documents of the Licensed Software 5.19 Quantification in units μCi/Bq Quality assurance feature to monitor and record parameters such as peak centroid, FWHM, Nuclide activities, MDA and other important results. The QA data stored and the results obtained should be displayed in a chart type or as a report template. 5.21 Ability to review, adjust, or add/delete peaks in specific regions/ROI 5.22 Python scripting integration for customizable results. Results review and reporting in software Should be organized into multiple sectioned based pages Should provide information such as basic sample information; a spectral image with count 			on	
 5.18 Hard Copy and complete documents of the Licensed Software 5.19 Quantification in units μCi/Bq Quality assurance feature to monitor and record parameters such as peak centroid, FWHM, Nuclide activities, MDA and other important results. The QA data stored and the results obtained should be displayed in a chart type or as a report template. 5.21 Ability to review, adjust, or add/delete peaks in specific regions/ROI 5.22 Python scripting integration for customizable results. Results review and reporting in software 5.23 Should be organized into multiple sectioned based pages Should provide information such as basic sample information; a spectral image with count 				
 Quantification in units μCi/Bq Quality assurance feature to monitor and record parameters such as peak centroid, FWHM, Nuclide activities, MDA and other important results.				
Quality assurance feature to monitor and record parameters such as peak centroid, FWHM, Nuclide activities, MDA and other important results. The QA data stored and the results obtained should be displayed in a chart type or as a report template. 5.21 Ability to review, adjust, or add/delete peaks in specific regions/ROI 5.22 Python scripting integration for customizable results. Results review and reporting in software 5.23 Should be organized into multiple sectioned based pages Should provide information such as basic sample information; a spectral image with count			censed Software	
Nuclide activities, MDA and other important results. The QA data stored and the results obtained should be displayed in a chart type or as a report template. 5.21 Ability to review, adjust, or add/delete peaks in specific regions/ROI 5.22 Python scripting integration for customizable results. - Results review and reporting in software 5.23 Should be organized into multiple sectioned based pages Should provide information such as basic sample information; a spectral image with count	5.19			
The QA data stored and the results obtained should be displayed in a chart type or as a report template. 5.21 Ability to review, adjust, or add/delete peaks in specific regions/ROI 5.22 Python scripting integration for customizable results. - Results review and reporting in software 5.23 Should be organized into multiple sectioned based pages 5.24 Should provide information such as basic sample information; a spectral image with count				
The QA data stored and the results obtained should be displayed in a chart type or as a report template. 5.21 Ability to review, adjust, or add/delete peaks in specific regions/ROI 5.22 Python scripting integration for customizable results. - Results review and reporting in software 5.23 Should be organized into multiple sectioned based pages 5.24 Should provide information such as basic sample information; a spectral image with count	5 20			
5.21 Ability to review, adjust, or add/delete peaks in specific regions/ROI 5.22 Python scripting integration for customizable results. - Results review and reporting in software 5.23 Should be organized into multiple sectioned based pages Should provide information such as basic sample information; a spectral image with count	0.20		nould be displayed in a chart type or as a report	
5.22 Python scripting integration for customizable results. - Results review and reporting in software 5.23 Should be organized into multiple sectioned based pages Should provide information such as basic sample information; a spectral image with count				
- Results review and reporting in software - 5.23 Should be organized into multiple sectioned based pages 5 Should provide information such as basic sample information; a spectral image with count		, ,	1 0	
5.23 Should be organized into multiple sectioned based pages Should provide information such as basic sample information; a spectral image with count	5.22		results.	
Should provide information such as basic sample information; a spectral image with count	-	Results review and reporting in software		-
	5.23			
information in one section	5 24	1	ple information; a spectral image with count	
5.25 Should provide information such as MCA parameters, library information, analysis algorithm	5.25	Should provide information such as MCA para	ameters, library information, analysis algorithm	

	annested details at a in west costion	
	executed details etc. in next section.	
5.26	Should provide ROI, FWHM, peak area, uncertainties and background counts information,	
	with an option to sort the results using header in next section.	
	Nuclide Results section: Should provide summary of nuclide activities, MDA results for	
5.27	identified and library radionuclides, Uncertainties. Provision to expand the radionuclide to	
	view the results of each emission line. Provision to toggle the uncertainty results from	
	percentage and absolute values.	
5.28	Provision to inspect any emission line in the results page by clicking on the energy and can	
	launch Interactive peak fit for reviewing the peaks. Provision for flagging the energy line that needs to be reviewed later stages for perfect	
5.29	analysis.	
6.00	Data Acquisition System	1 Number
6.01	Desktop computer preloaded with licensed latest Windows OS.	1 Number
6.02		
6.02	The computer shall have communication port to communicate with the instrument	
6.03	Core i7-14700K or greater processor, 32GB DDR4 RAM, 1TB SSD, 24" LED display unit, Latest	
7.00	MS Office, wireless keyboard & optical mouse. Lead Shield for the HPGe system	1 Number
7.00	Lead shield for lower background with a support table suitable for the detector as detailed	1 Number
7.01	Shielded Cavity Dimensions:	
7.02	J and the state of	
7.02	≥250 mm gap between top of the detector and sliding door ≥150 mm gap between detector well and shielding inner well	
7.02	>150 mm gap between detector wall and shielding inner wall Chielding Torres Colid goest recognized the denith sheel goesing and denited 7 lines.	
7.03	Shielding Type: Solid-cast reprocessed lead with steel casing and graded-Z liner.	
	Shielding material Specifications:	
	• 12.7 mm (0.5 in.) low-carbon steel casing	
7.04	• 101 mm (4 in.) reprocessed lead	
	• 0.5 mm (0.02 in.) tin sheet liner	
	• 1.6 mm (0.064 in.) soft-copper sheet liner	
	Support Stand Material, Low-carbon steel square tubing and plate. OFF (OFF) (a standard plate)	
8.00	OEM/OEM authorization letter should be provided along with the offer. Otherwise, offer	
	will be rejected technically.	
0.00	All the parameters mentioned as complied should be provided with sufficient documentary	
9.00	evidence such as OEM datasheet or on OEM letterhead, along with the offer. Without	
	sufficient documentary evidence, offers will be rejected technically.	
	Warranty	
	System should be supplied with two years comprehensive warranty from the date of installation (inclusive of all parts of LIB Co. and accessories, LIBC Patteries etc.)	
	installation (inclusive of all parts of HP Ge and accessories, UPS, UPS Batteries etc).	
	The successful bidder will be responsible for the comprehensive maintenance of the	
	supplied machine and its related accessories free of charge during the warranty	
10.00	period. In case of any complaint in the working of the instrument during the warranty.	
10.00	• In case of any complaint in the working of the instrument during the warranty	
	period, the firm shall attend the same within 48 hours from the time of logging of 1st complaint call by IREL and the instrument must be repaired within 10 days	
	thereafter. Further, in case of failure to do so, penalty @ Rs. 2000/- per day will be	
	applicable.	
	Installation & Training	
	Supplier should provide training on operation and application of the equipment.	
	Installation and commissioning to be done at our site with full demonstration of	
	software and day to day maintenance of the equipment. One week comprehensive	
11.00	training on the use of the systems should be provided by the Supplier at our site at	
11.00	IRE, RED, Udyogamandal after successful installation and commissioning. During	
	the training, the supplier should provide necessary inputs through the engagement of	
	application executive for development of methods for the analysis of all Rare earth	
	products as specified by IRE, RED.	
<u> </u>	products as specifical by Italy Italb.	

	 Transportation, boarding and lodging for the trainers will be borne by the supplier during the training at IRE, RED. Qualifications for installation crew to meet the related statutory body requirements and approvals shall also be complied 	
12.00	 Other requirements All necessary pre-installation requisites for complete installation of equipment will be carried out by the Supplier. Power requirement for the system and any local ancillary equipment/ instrument necessary to run the system, in addition to the above, shall be informed to IREL in advance, before supplying the instrument. The successful bidder has to give undertaking stating capability to supply the required spares and optical components for a period of minimum 10 years after completion of two years comprehensive warranty period as per requirement ensuring instrument operating in good condition and able to carry-out AMC after warranty. A maintenance chart for all the components in the system shall be provided. Manual: One set of Operating manual and Service manual (in English language) with complete circuit diagrams should be provided with the instrument. The Manual should be presented in both, hard and soft copy. 	
13.00	Pre-requisites All necessary installation requirements to be informed before supplying the instrument.	
14.00	IREL scope of supply Power connections, Bench/platform for installing the equipment, water & electricity required for commissioning.	

$\frac{\text{GENERAL CONDITIONS OF CONTRACT (GCOC)}}{\underline{\text{INDEX}}}$

Sl. No.	Particulars
1	DEFINITIONS
2	SELLER TO INFORM
3	CONFLICT AND INTERPRETATION OF DOCUMENTS
4	COUNTRY OF ORIGIN
5	SCOPE OF CONTRACT
6	STANDARDS
7	INSTRUCTIONS, DIRECTION & CORRESPONDENCE,
	CONTRACT OBLIGATIONS
8	MODIFICATION IN CONTRACT
9	PATENT RIGHTS, LIABILITY & COMPLIANCE OF
	REGULATIONS
10	INSPECTION, TESTING & EXPEDITING
11	TIME SCHEDULE & PROGRESS REPORTING
12	DELIVERY & DOCUMENTS
13	TRANSIT RISK INSURANCE
14	TRANSPORTATION
15	INCIDENTAL SERVICES
16	SPARE PARTS, MAINTENANCE TOOLS, LUBRICANTS
17	GUARANTEE
18	PRICES
19	SUBLETTING AND ASSIGNMENT
20	TIME AS ESSENCE OF CONTRACT
21	DELAYS IN THE SELLER'S PERFORMANCE
22	LIQUIDATED DAMAGES FOR DELAYED DELIVERY
23	REJECTIONS, REMOVAL OF REJECTED EQUIPMENT &
	REPLACEMENT
24	TERMINATION FOR DEFAULT
25	FORCE MAJEURE
26	RESOLUTION OF DISPUTES/ ARBITRATION
27	TAXES & DUTIES
28	PERMITS & CERTIFICATES
29	FALL CLAUSE
30	LIMITATION OF LIABILIT
31	METHOD OF BLACKLISTING VENDORS
32	SECRECY
33	GENERAL
34	RISK PURCHASE CLAUSE

General Conditions of Contract (GCC-Goods)

Definitions In this document, General Conditions of Contract (GCC-Goods), the following terms shall have the following respective meanings:

- BIDDER: Designates the individual or legal entity which has made a proposal, a tender or a bid with the aim of concluding a Contract with the PURCHASER. 1.0
- 1.1
- 1.2 CONTRACT shall mean Purchase Order/ Service Contracts and all attached exhibits and documents referred to therein and all terms and conditions thereof together with any subsequent modifications thereto.
- 1.3
- CONTRACT PRICE shall mean the price payable to the Seller under the Contract for the full and proper performance of his contractual obligations.

 COMPLETION DATE shall mean the date on which the goods are successfully supplied/commissioned by the Seller and handed over to the PURCHASER. 1.4
- 1.5 COMMERCIAL OPERATION shall mean the condition of the operation in which the complete equipment covered under the Contract is officially declared by the PURCHASER to be available for continuous operation at different loads up to and including rated capacity.

 DELIVERY terms shall be interpreted as per INCO TERMS 2000 in case of Contract with a foreign Bidder and as the Goods-in-ward date in the case of a contract with an Indian Bidder.
- 16 1.7 DRAWINGS shall mean and include Engineering drawings, sketches showing plans, sections and elevations in relation to the Contract together with modifications and/or revisions
- thereto. OFFICER-IN-CHARGE(OIC)/ENGINEER-IN-CHARGE(EIC): OFFICER-IN-CHARGE (OIC) / Engineer-in-Charge of the Project SITE shall mean the person designated from time to time 1.8
 - by PURCHASER/CONSULTANT at SITE and shall include those who are expressly authorized by him/her to act for and on his/her behalf for operation of this CONTRACT. The Officer/ Engineer-in-charge shall have authority for
 - \bullet General supervision, Follow up of supply and direction of the work
 - direction to stop the work whenever such stoppage may be necessary to ensure the proper execution of the Contract
 to reject all works and materials which do not conform to the contract The OIC/EIC shall have neither any authority to relieve the contractor of any of his duties / obligations under the contract nor ordering any work involving delay or any extra payment by IREL or making any variation of or in the work except otherwise expressly provided here-in-under or elsewhere in the contract.
- FINAL ACCEPTANCE shall mean the PURCHASER's written acceptance of the Works performed under the Contract after successful completion of performance and guarantee test. 1.9
- 1.10
- GOODS shall mean articles, materials, equipment, design and drawings, data and other property to be supplied by Seller to complete the contract.

 INSPECTOR shall mean any person or outside Agency nominated by PURCHASER/CONSULTANT through CONSULTANT to inspect equipment, stagewise as well as final, before dispatch, at SELLER's works and on receipt at SITE as per terms of the CONTRACT. 1.11
- 1.12 INITIAL OPERATION shall mean the first integral operation of the complete equipment covered under the Contract with sub-systems and supporting equipment in service or available for service.
- PURCHASER shall mean IREL (India) Limited (IREL) having its registered office at Plot No. 1207, Veer Savarkar Marg, Near Siddhi Vinayak Temple, Prabhadevi, Mumbai -400028. The 1 13 term PURCHASER includes successors, assigns of IREL.
- PERFORMANCE AND GUARANTEE TESTS shall mean all operational checks and tests required to determine and demonstrate capacity, efficiency and operating characteristics as 1.14 specified in the Contract documents.
- PROJECT designates the aggregate of the Goods and/or Services to be provided by one or more Contractors. 1.15
- QUANTITIES Bills of quantities designate the quantity calculations to be taken into account when these calculations are made from detailed or construction drawings, or from work 1.16 actually performed, and presented according to a jointly agreed breakdown of the Goods and/or Services.
- SELLER shall mean the person, firm or company with whom PURCHASE ORDER/ SERVICE CONTRACT is placed/ entered into by PURCHASER for supply of equipment, materials 1.17 and services. The term Seller includes its successors and assigns.
- SERVICE shall mean erection, installation, testing, commissioning, provision of technical assistance, training and other such obligations of the Seller covered under the Contract.
- SITE designates the land and/or any other premises on, under, in or across which the Goods and/or Services have to be supplied, erected, assembled, adjusted, arranged and/or 1.19 commissioned.
- SPECIFICATIONS shall mean and include schedules, details, description, statement of technical data, performance characteristics, standards (Indian as well as International) as applicable 1.20 and specified in the Contract.
- 1.21 SUB-CONTRACT shall mean order placed by the Seller, for any portion of the contracted work, after necessary consent and approval of PURCHASER.
- SUB-CONTRACTOR shall mean the person named in the CONTRACT for any part of the work or any person to whom any part of the CONTRACT has been sub-let by the SELLER with 1.22 the consent in writing of the CONSULTANT/PURCHASER and will include the legal representatives, successors, and permitted assigns of such person.
- 1 23 START-UP shall mean the time period required to bring the equipments covered under the Contract from an inactive condition, when construction is essentially complete to the state of readiness for trial operation. The start-up period shall include preliminary inspection and check out of equipment and supporting subsystems, initial operation of the complete equipments covered under the Contract to obtain necessary pre-trial operation data, perform calibration and corrective action, shutdown inspection and adjustment prior to the trial operation period.
- TESTS shall mean such process or processes to be carried out by the Seller as are prescribed in the Contract or considered necessary by PURCHASER or his representative in order to 1.24 ascertain quality, workmanship, performance and efficiency of equipment or part thereof.
- 1 25 TESTS ON COMPLETION shall mean such tests as prescribed in the Contract to be performed by the Seller before the Works are taken over by the PURCHASER.

2.0 SELLER TO INFORM:

The Seller shall be deemed to have carefully examined all contract documents to his entire satisfaction. Any lack of information shall not in any way relieve the Seller of his responsibility to fulfil his obligation under the Contract

CONFLICT AND INTERPRETATION OF DOCUMENTS: 3.0

- Words imparting the singular only also include the plural and vice versa where the context requires so; words imparting persons include firms or corporations and vice versa where the 3.1 context requires.
- 3.2 Word imparting masculine gender includes the feminine gender and vice versa where the context requires so.
- The several documents forming the contract are to be taken as mutually explanatory of one another. 3.3
- In case conflicting statements or directives occur among the contract documents, it shall be the responsibility of the successful bidder to notify IREL, with a copy sent concurrently to the 34 OIC/EIC, immediately in writing and obtain instructions from IREL to eliminate the conflict.
- 3.5 The successful bidder shall notify IREL, with a copy sent concurrently to the EIC/OIC, promptly of any discrepancies, omissions or doubts it may have, regarding drawings, specifications or other documents. Noted or calculated dimensions shall always be followed
- In the event of conflict between various documents forming the contract, the relevant terms and conditions of the purchase order shall prevail over those of all such other documents 3.6 forming the contract and binding on the contractor and IREL.
- 3.7
- In case of any conflict between the General Conditions of contract and the Special Conditions of the contract, the Special Conditions of contract shall prevail.

 Country of Origin: For purposes of this Clause "Origin" means the place where the Goods were produced, or from which the services are supplied. Goods are produced when, through manufacturing, processing or substantial and major assembling of components, a commercially recognized new product results that is substantially different in basic characteristics or in purpose or utility from its components. 4.0

5.0

- Scope of the CONTRACT shall be as defined in the PURCHASE ORDER/CONTRACT specifications, drawings and Annexure thereto. 5.1
- Completeness of the EQUIPMENT shall be the responsibility of the SELLER. Any equipment, fittings and accessories which may not be specifically mentioned in the specifications or drawings, but which are usual or necessary for the satisfactory functioning of the equipment (successful operation and functioning of the EQUIPMENT being SELLER's responsibility) 5.2 shall be provided by SELLER without any extra cost. 5.3
- The SELLER shall follow the best modern practices in the manufacture of high grade EQUIPMENT notwithstanding any omission in the specifications. The true intent and meaning of these documents is that SELLER shall in all respects, design, engineer, manufacture and supply the equipment in a thorough workmanlike manner and supply the same in prescribed time to the entire satisfaction of PURCHASER.
- The SELLER shall furnish minimum three (03) copies in English language of Technical documents, final drawings, preservation instructions, operation and maintenance manuals, test 5.4 certificates, spare parts catalogues for all equipment to the PURCHASER.
- 5.5 The documents once submitted by the SELLER shall be firm and final and not subject to subsequent changes. The SELLER shall be responsible for any loss to the PURCHASER/CONSULTANT consequent to furnishing of incorrect data/drawings.
- All dimensions and weight should be in metric system.
- All equipment to be supplied and work to be carried out under the CONTRACT shall conform to and comply with the provisions of relevant regulations/Acts (State Government or 5.7 Central Government) as may be applicable to the type of equipment/work carried out and necessary certificates shall be furnished.
- The Seller shall provide cross sectional drawings, wherever applicable, to identify the spare part numbers and their location. The size of bearings, their make and number shall be 5.8 furnished by the SELLER.
- Specifications, design and drawings issued to the SELLER along with NIT/RFQ and CONTRACT are not sold or given but loaned. These remain property of PURCHASER/CONSULTANT or its assigns and are subject to recall by PURCHASER/CONSULTANT. The SELLER and his employees shall not make use of the drawings, specifications and technical information for any purpose at any time except for manufacture against the CONTRACT and shall not disclose the same to any person, firm or corporate body, without written permission of PURCHASER/CONSULTANT. All such details shall be kept confidential. 5.9

5.10 SELLER shall pack, protect, mark and arrange for despatch of EQUIPMENT as per instructions given in the CONTRACT.

6.0

7.1

8.1

9.0

9.1

9.2

10.2

10.3

10.4

10.6 10.7

10.8

10.9

11.2

The GOODS supplied under the CONTRACT shall conform to the standards mentioned in the Technical Specifications, or such other standards which ensure equal or higher quality, and when no applicable standard is mentioned, to the authoritative standard appropriate to the GOODS' country of origin and such standards shall be the latest issued by the concerned

7.0

- Instructions, Direction & Correspondence
 The materials described in the CONTRACT are to be supplied according to the standards, data sheets, tables, specifications and drawings attached thereto and/or enclosed with the CONTRACT, itself and according to all conditions, both general and specific enclosed with the contract, unless any or all of them have been modified or cancelled in writing as a whole or
 - a. All instructions and orders to SELLER shall, excepting what is herein provided, be given by PURCHASER/CONSULTANT. b. All the work shall be carried out under the direction of and to the satisfaction of PURCHASER/CONSULTANT.

 - c. All communications including technical/commercial clarifications and/or comments shall be addressed to CONSULTANT in quintuplicate with a copy to PURCHASER and shall always bear reference to the CONTRACT.
 - d. Invoices for payment against CONTRACT shall be addressed to PURCHASER. The CONTRACT number shall be shown on all invoices, communications, packing lists, containers and bills of lading, etc.
 - e. Invoice must contain seller's GST Registration number, PAN, Bank detail of supplier, GST Registration number of IREL Factory/ Office, HSN Code for products being supplied. Supplier must submit e-way bill, e-invoice from GSTM Portal.

Contract Obligations:

- 7.2.1 If after award of the contract, the Seller does not acknowledge the receipt of award or fails to furnish the performance guarantee within the prescribed time limit, the PURCHASER reserves the right to cancel the contract and apply all remedies available to him under the terms and conditions of this contract.
- Once a contract is confirmed and signed, the terms and conditions contained therein shall take precedence over the Seller's bid and all previous correspondence. 7.2.2

8.0 Modification in Contract:

All modifications leading to changes in the CONTRACT with respect to technical and/or commercial aspects including terms of delivery, shall be considered valid only when accepted in writing by PURCHASER/CONSULTANT by issuing amendment to the CONTRACT. Issuance of acceptance or otherwise in such cases shall not be any ground for extension of agreed

delivery date and also shall not affect the performance of contract in any manner except to the extent mutually agreed through a modification of contract.

PURCHASER/CONSULTANT shall not be bound by any printed conditions or provisions in the SELLER's Bid Forms or acknowledgment of CONTRACT, invoices, packing list and 8.2 other documents which purport to impose any conditions at variance with or be supplemental to CONTRACT.

Patent Rights, Liability & Compliance of Regulations:

SELLER hereby warrants that the use or sale of the materials delivered hereunder will not infringe claims of any patent covering such material and SELLER agrees to be responsible for and to defend at his sole expense all suits and proceedings against PURCHASER based on any such alleged patent infringement and to pay all costs, expenses and damages which PURCHASER and/or CONSULTANT may have to pay or incur by reason of any such suit or proceedings.

The SELLER shall indemnify the PURCHASER against all third party claims of infringement of patent, trade mark or industrial design rights arising from use of the GOODS or any part

- thereof in the PURCHASER's country.
- SELLER shall also protect and fully indemnify the PURCHASER from any claims from SELLER'S workmen/employees or their heirs, dependants, representatives, etc. or from any other person/persons or bodies/companies etc. for any acts of commissions or omission while executing the CONTRACT. 9.3
- SELLER shall be responsible for compliance with all requirements under the laws and shall protect and indemnify completely the PURCHASER from any claims/penalties arising out of 9.4 any infringements. 9.5
 - PURCHASER will promptly notify the SELLER in writing of any such claim, suit, action or proceeding coming to its attention, giving authority and all available information and assistance for the SELLER's defence of the same. If at any time the installation of the plant or any part thereof, or the use thereof in India for the purpose for which it is furnished or the sale of products produced therewith, is prevented or enjoyed because of patent infringement or claimed infringement, the SELLER shall promptly at his own expense, either procure for PURCHASER the rights to use and continue to use such plant or replace the same at his own expense with equally efficient non-infringing plant satisfactory under all requirements of the contract, so that the operation of PURCHASER's plant will not be unduly delayed or interrupted. If shipment of the plant, or any part thereof is prevented by attachment, injunction or otherwise, or in the course of transit from the SELLER's factory or other point of origin to the site of PURCHASER, as a result of any claim of patent infringement the SELLER shall, at his own cost and expense, promptly furnish and post the necessary bond or take such other steps as may be necessary to enable shipment to be made without delay. PURCHASER will have the right to retain counsel of its own choice to collaborate in the defence of any such claim, suit action or proceeding.

10.0 Inspection, Testing & Expediting 10.1

The PURCHASER or its representative shall have the right to inspect and/or to test the GOODS to confirm their conformity to the CONTRACT specifications. The special conditions of CONTRACT and/or the Technical Specifications shall specify what inspections and tests the PURCHASER requires and where they are to be conducted. The PURCHASER shall notify the SELLER in writing the identity of any representative(s) retained for these purposes.

The inspections and tests may be conducted on the premises of the SELLER or his sub-contractor(s), at point of DELIVERY and/or at the GOODS' final destination, When conducted on the premises of the SELLER or his sub-contractor (s), all reasonable facilities and assistance including access to the drawings and production data shall be furnished to the inspectors at no charge to the PURCHASER.

Should any inspected or tested GOODS fail to conform to the specifications, the PURCHASER may reject them and the SELLER shall either replace the rejected GOODS or make all alterations necessary to meet Specifications' requirements, free of cost to the PURCHASER.

The PURCHASER's right to inspect, test and where necessary reject the GOODS after the GOODS' arrival in the PURCHASER's country shall in no way be limited or waived by reason of the GOODS having previously been inspected, tested and passed by the PURCHASER, or their representative prior to the GOODS shipment from the country of origin.

The INSPECTOR shall follow the progress of the manufacture of the GOODS under the CONTRACT to ensure that the requirements outlined in the CONTRACT are not being deviated with respect to schedule and quality.

SELLER shall allow the INSPECTOR to visit, during working hours, the workshops relevant for execution of the CONTRACT during the entire period of CONTRACT validity.

In order to enable PURCHASER's representatives to carryout the inspection in time, SELLER shall notify PURCHASER 15 days before assembly, testing and packing of main

EQUIPMENT. If requested, SELLER shall assist PURCHASER's representatives in getting visas in the shortest possible time (applicable only in case of foreign order).

SELLER shall place at the disposal of the INSPECTOR, free of charge, all tools, instruments, and other apparatus necessary for the inspection and/or testing of the GOODS. The INSPECTOR is entitled to prohibit the use and dispatch of GOODS and/or materials which have failed to comply with the characteristics required for the GOODS during tests and inspections.

SELLER shall advise in writing of any delay in the inspection program at the earliest, describing in detail the reasons for delay and the proposed corrective action.

ALL TESTS and trials in general, including those to be carried out for materials not manufactured by SELLER shall be witnessed by the INSPECTOR. Therefore, SELLER shall confirm to 10.10 PURCHASER by fax or e-mail about the exact date of inspection with at least 30 days notice. SELLER shall specify the GOODS and quantities ready for testing and indicate whether a preliminary or final test is to be carried out. 10 11

If on receipt of this notice, should PURCHASER waive the right to witness the test, timely information will be given accordingly.

Any and all expenses incurred in connection with tests, preparation of reports and analysis made by qualified laboratories, necessary technical documents, testing documents and drawings shall be at SELLER's cost. The technical documents shall include the reference and numbers of the standards used in the construction and, wherever deemed practical by the 10.12 INSPECTOR, copy of such standards.

10.13 Nothing in Article-10 shall in any way release the SELLER from any warranty or other obligations under this CONTRACT. 10.14

Arrangements for all inspections required by Indian Statutory Authorities and as specified in technical specifications shall be made by SELLER.

10.15 Inspection & Rejection of Materials by consignees:

When materials are rejected by the consignee, the supplier shall be intimated with the details of such rejected materials, as well as the reasons for their rejection, also giving location where such materials are lying at the risk and cost of the supplier. The supplier will be called upon either to remove the materials or to give instructions as to their disposal within 14 days and in the case of dangerous, infected and perishable materials within 48 hours, failing which the consignee will either return the materials to the SELLER freight to pay or otherwise dispose them off at the SELLER's risk and cost. The PURCHASER shall also be entitled to recover handling and storage charges for the period, during which the rejected materials are not removed @ 5% of the value of materials for each month or part of a month till the rejected materials are finally disposed off.

11.0 Time Schedule & Progress Reporting

Together with the Contract confirmation, SELLER shall submit to PURCHASER, his time schedule regarding the documentation, manufacture, testing, supply, erection and 11.1 commissioning of the GOODS

PURCHASER's/CONSULTANT's representatives shall have the right to inspect SELLER's premises with a view to evaluating the actual progress of work on the basis of SELLER's time schedule documentation

11.3 Irrespective of such inspection, SELLER shall advise PURCHASER at the earliest possible date of any anticipated delay in the progress.

Notwithstanding the above, in case progress on the execution of contract at various stages is not as per phased time schedule and is not satisfactory in the opinion of the PURCHASER/CONSULTANT which shall be conclusive or SELLER shall neglect to execute the CONTRACT with due diligence and expedition or shall contravene the provisions of the 11.4 CONTRACT, PURCHASER/CONSULTANT may give notice of the same in writing to the SELLER calling upon him to make good the failure, neglect or contravention complained of. Should SELLER fail to comply with such notice within the period considered reasonable by PURCHASER/CONSULTANT, the PURCHASER/ CONSULTANT shall have the option and be at liberty to take the CONTRACT wholly or in part out of the SELLER's hand and make alternative arrangements to obtain the requirements and completion of CONTRACT at the SELLER's risk and cost and recover from the SELLER, all extra cost incurred by the PURCHASER on this account. In such event, PURCHASER / CONSULTANT shall not be responsible for any loss that the SELLER may incur, and SELLER shall not be entitled to any gain. PURCHASER/CONSULTANT shall, in addition, have the right to encash Performance Guarantee in full or part.

12.0 Delivery & Documents:

- 12.1 Delivery of the GOODS shall be made by the SELLER in accordance with terms specified in the CONTRACT, and the goods shall remain at the risk of the SELLER until delivery has been completed.
- 12.2 Delivery shall be deemed to have been made:
 - a) In the case of FOB, CFR & CIF Contracts, when the Goods have been put on board the ship, at the specified port of loading and a clean Bill of Lading is obtained. The date of Bill of Lading shall be considered as the delivery date.
 - b) In case of FOT despatch point contract (For Indian bidder), on evidence that the goods have been loaded on the carrier and a negotiable copy of the GOODS receipt obtained, the date of LR/GR shall be considered as the date of delivery. c) In case of FOT site/ FOR Destination (for Indian bidders), on receipt of goods by PURCHASER/Consultant at the designated site(s).
- The delivery terms are binding and essential and consequently, no delay is allowed without the written approval of PURCHASER/CONSULTANT. Any request concerning delay will be void unless accepted by PURCHASER/CONSULTANT through a modification to the CONTRACT. 12.3
- Delivery time shall include time for submission of drawings for approval, incorporation of comments, if any, and final approval of drawings by PURCHASER/CONSULTANT 12.4
- 12.5 In the event of delay in delivery, Liquidated Damages as stipulated in Article - 22.1 shall apply.
- 12.6 The documentation, in English Language, shall be delivered in due time, in proper form and in the required number of copies as specified in the contract.
- The additional copies of final drawings and instructions will be included in the package of goods, properly enveloped and protected. The SELLER should comply with the Packing, Marking and Shipping Documentation Specifications enclosed. 12.7
- 12.8

13.0 Transit Risk Insurance

- All goods supplied under the contract shall be fully insured in a freely convertible currency against loss or damage incidental to manufacture or acquisition, transportation, storage and 13.1 delivery.
 - Where delivery is on FOB or CFR basis, marine insurance shall be the responsibility of the Purchaser. Insurance Requirements > Indigenous Bidders: Transit risk insurance from F.O.T. despatch point onwards shall be arranged and borne by PURCHASER. In case of F.O.T site contracts, transit insurance to be borne by SELLER. > Foreign Bidders: Marine insurance as well as transit insurance in Purchaser's country shall be arranged and borne by PURCHASER. The SELLER shall ensure that in effecting despatch of materials, the primary responsibility of the carriers for safe movement is always retained so that the PURCHASER's interests are fully safeguarded and are in no way jeopardised. The Seller shall furnish the cost of materials against each equipment.

Transportation

13.2

- Where the SELLER is required under the CONTRACT to deliver the GOODS FOB, transport of the GOODS until delivery, that is, upto and including the point of putting the GOODS on 14.1
- board the export conveyance at the specified port of loading, shall be arranged and paid for by the SELLER and the cost thereof shall be included in the Contract price.

 Where the SELLER is required under the CONTRACT to deliver the GOODS CFR or CIF, transport of the Goods to the port of discharge or such other point in the country of destination 14.2 as shall be specified in the CONTRACT shall be arranged and paid for by the SELLER and the cost thereof shall be included in the Contract price.

15.0

- 15.1 The Seller may be required to provide any or all of the following services:
- 15 1 1 Performance or supervision of onsite assembly and/or start-up of the supplied Goods:
- Furnishing tools required for assembly and/or maintenance of the supplied Goods: 15.1.2
- Performance or supervision or maintenance and/or repair of the supplied Goods, for a period of time agreed by the parties, provided that this service shall not relieve the Seller of any 15.1.3 warranty/ guarantee obligations under the Contract.
- Training of the Purchaser's personnel at the Seller's plant and/or at Site, in assembly, start-up operation, maintenance and/or repair of the supplied Goods at no extra cost. However, 15.1.4 Purchaser will bear boarding, lodging & personal expenses of Trainees
- Prices charged by the Seller for the preceding incidental services, shall not exceed the prevailing rates charged to other parties by the Seller for similar services. 15.2
- 15.3 When required, Seller shall depute necessary personnel for supervision and/or erection of the Equipment at site for duration to be specified by Purchaser on mutually agreed terms. Seller's personnel shall be available at Site within seven days for emergency action and twenty one days for medium and long-term assistance, from the date of notice given by Purchaser.
- The cost of incidental services shall not be included in the quoted prices. The cost of applicable incidental services should be shown separately in the price schedules whenever asked for 15.4

16.0 Spare Parts, Maintenance Tools, Lubricants

- Seller may be required to provide any or all of the following materials and notification pertaining to spare parts manufactured or distributed by the Seller. 16.1
- Such spare parts as the Purchaser may opt to purchase from the Seller, provided that his option shall not relieve the Seller of any warranty obligations under the Contract, and 16.1.2 In 16.1.1 the event of termination of production of the spare parts:
 - i) Advance notification to the Purchaser of the pending termination, in sufficient time to permit the Purchaser to procure needed requirements, and
 - ii) Following such termination, furnishing at no cost to the Purchaser, the blueprints, drawings and specifications of the spare parts, if any when, requested.
- Seller shall supply item wise list with value of each item of spare parts and maintenance tools requirements, along with full details of manufacturers/vendors for such 16.2 spares/maintenance tools for:
- 16.2.1 The construction, execution and commissioning.
- 16.2.2 Two years operation and maintenance. 16.3
 - Spare parts shall be new and of first-class quality as per engineering standards/ codes, free of any defects (even concealed), deficiency in design, materials and workmanship and also shall be completely interchangeable with the corresponding parts.
- 16.4 Type and sizes of bearings shall be clearly indicated. 16.5
 - Spare parts shall be packed for long storage under tropical climatic conditions in suitable cases, clearly marked as to intended purpose.
- A list of special tools and gauges required for normal maintenance and special handling and lifting appliances, if any, for the Goods shall be submitted to Purchaser. 16.6
- 16.7 Bidders should note that if they do not comply with Clause 16.2 above, their quotation may be rejected.
- 16.8
- Whenever lubricants are required, Seller shall indicate the quantity of lubricants required for the first filling, the frequency of changing, the quantity of lubricants required for the one 16.8.1 year's continuous operation and the types of recommended lubricants indicating the commercial name (trade-mark), quality and grade.
 - If Seller is unable to recommend specific oil, basic recommended characteristics of the lubricants shall be given.
- Seller shall indicate various equivalent lubricants available in India. 16.8.3

17.0

17.1

All Goods or Materials shall be supplied strictly in accordance with the specifications, drawings, data sheets, other attachments and conditions stated in the Contract. No deviation from such specifications or alterations or of these conditions shall be made without PURCHASER'S / CONSULTANT'S agreement in writing which must be obtained before any work against the order is commenced. All materials supplied by the SELLER pursuant to the Contract (irrespective of whether engineering, design data or other information has been furnished, reviewed or approved by PURCHASER) are guaranteed to be of the best quality of their respective kinds (unless otherwise specifically authorised in writing by PURCHASER) and shall be free from faulty design, workmanship and materials, and to be of sufficient size and capacity and of proper materials so as to fulfil in all respects all operating conditions, if any, specified in the Contract.

If any trouble or defect, originating with the design, material, workmanship or operating characteristics of any materials, arises at any time prior to twelve(12) months from the date of the first commercial operation of the Plant for which the materials supplied under the Contract form a part thereof, or Fifteen (15) months from the date of last shipment whichever period shall first expire, and the SELLER is notified thereof, SELLER shall, at his own expense and as promptly as possible, make such alterations, repairs and replacements as may be necessary to permit the materials to function in accordance with the specifications and to fulfil the foregoing guarantees. PURCHASER may, at his option, remove such defective materials, at SELLER'S expense in which event SELLER shall, without cost to PURCHASER and as promptly as possible, furnish and install proper materials. Repaired or replaced materials shall be similarly guaranteed by the SELLER for a period of no less than twelve (12) months from the date of replacement/repair.

In the event that the materials supplied do not meet the specifications and/or not in accordance with the drawings, data sheets or the terms of the Contract and rectification is required at site, PURCHASER shall notify the SELLER giving full details of differences. The SELLER shall attend the site within seven (7) days of receipt of such notice to meet and agree with

representatives of PURCHASER, for taking the action required to correct the deficiency. Should the SELLER fail to attend meeting at Site within the time specified above, PURCHASER shall immediately rectify the work/ materials and SELLER shall reimburse PURCHASER all costs and expenses incurred in connection with such trouble or defect.

PERFORMANCE GUARANTEE OF EQUIPMENT 17.2

- 1721 SELLER shall guarantee that the performance of the EQUIPMENT supplied under the CONTRACT shall be strictly in conformity with the specifications and shall perform the functions specified under the CONTRACT.
- 17.2.2 If the SELLER fails to prove the guaranteed performance of the EQUIPMENT set forth in the specification, the SELLER shall investigate the causes and carry out necessary rectifications/modifications to achieve the guaranteed performance. In case the SELLER fails to do so within a reasonable period, the SELLER shall replace the EQUIPMENT and prove guaranteed performance of the new equipment without any extra cost to PURCHASER.
- 17.2.3 If the SELLER fails to prove the guarantee within a reasonable period, PURCHASER/CONSULTANT shall have the option to take over the EQUIPMENT and rectify, if possible, the EQUIPMENT to fulfil the guarantees and/or to make necessary additions to make up the deficiency at Seller's risk and cost. All expenditure incurred by the PURCHASER/ CONSULTANT in this regard shall be to SELLER's account.

18.0 Prices:

21.0

Prices charged by the SELLER for Goods delivered and services performed under the CONTRACT shall not, with the exception of any price adjustments authorized by the Contract vary from the prices quoted by the SELLER in his bid.

The price indicated in the Purchase Order is firm and not subject to alteration on any account unless specified otherwise in the terms and conditions of Purchase Order. However, in case of delayed delivery, price reduction clause shall apply.

19.0 Subletting and Assignment:

The SELLER shall not without previous consent in writing of the PURCHASER, sublet, transfer or assign the contract or any part thereof or interest therein or benefit or advantage thereof in any manner whatsoever. Provided, nevertheless, that any such consent shall not relieve the SELLER from any obligation, duty or responsibility under the contract.

20.0 Time as Essence of Contract:

The time and date of delivery/completion of the GOODS/SERVICES as stipulated in the Contract shall be deemed to be the essence of the Contract.

Delays in The Seller's Performance:

If the specified delivery schedule is not adhered to or the progress of manufacture or supply of the items is not satisfactory or is not in accordance with the progress schedule, the PURCHASER has the right to:

- hire for period of delay from elsewhere goods which in PURCHASER's opinion will meet the same purpose as the goods which are delayed and SELLER shall be liable without limitation for the hire charges;
- ii) cancel the CONTRACT in whole or in part without liability for cancellation charges. In that event, PURCHASER may procure from elsewhere ii) goods which PURCHASER's opinion would meet the same purpose as the goods for which CONTRACT is cancelled and SELLER shall be liable without limitations for the difference between the cost of such substitution and the price set forth in the CONTRACT for the goods involved; or iii) hire the substitute goods vide (I) above and if the ordered goods continue to remain undelivered thereafter, cancel the order in part or in full vide (ii) above.
- Any un-excusable delay by the SELLER or his subcontractor shall render the SELLER liable, without prejudice to any other terms of the Contract, to any or all of the following sanctions: 21.1 forfeiture of Contract performance guarantee, imposition of Liquidated Damages for delay in delivery, termination of the contract for default and purchase from another source at the risk and cost of the SELLER.
- 22.1 Liquidated Damages Schedule for Delayed Delivery. Subject to Article -25, if the SELLER fails to deliver any or all of the GOODS or performance the services within the time period (s) specified in the CONTRACT, the PURCHASER shall, without prejudice to his other remedies under the CONTRACT, deduct from the CONTRACT PRICE, a sum calculated on the basis of the CONTRACT PRICE, including subsequent modifications

22 1 1 Deductions shall apply as per following formula:

A sum equivalent to 0.5 (Half) per cent of the prices of any portion of stores delivered late, for each week or part thereof of delay. The total damages shall not exceed 10 (Ten) percent of the value of delayed goods. The LD cannot exceed the amount stipulated in the contract.

In case of delay in delivery on the part of SELLER, the invoice/document value shall be reduced proportionately for the delay and payment shall be released accordingly.

22.2 22.3

In the event the invoice value is not reduced proportionately for the delay, the PURCHASER may deduct the amount so payable by SELLER, from any amount falling due to the SELLER or by recovery against the Performance Guarantee. Both seller and PURCHASER agree that the above percentages of price reduction are genuine pre estimates of the loss/damage which the PURCHASER would have suffered on account of delay/breach on the part of the SELLER and the said amount will be payable on demand without there being any proof of the actual loss/or damage caused by such breach/delay. A decision of the PURCHASER in the matter of applicability of price reduction shall be final and binding.

23.0 Rejections, Removal of Rejected Equipment & Replacement 23.1

- Preliminary inspection at SELLER's works by INSPECTOR shall not prejudice PURCHASER's/ CONSULTANT's claim for rejection of the EQUIPMENT on final inspection at SITE or claims under warranty provisions.
- 23.2 If the EQUIPMENTS are not of specification or fail to perform specified functions or are otherwise not satisfactory the PURCHASER/CONSULTANT shall be entitled to reject the EQUIPMENT/MATERIAL or part thereof and ask free replacement within reasonable time failing which obtain his requirements from elsewhere at SELLER's cost and risk.
- Nothing in this clause shall be deemed to deprive the PURCHASER AND/OR AFFECT ANY rights under the Contract which it may otherwise have in respect of such defects or 23.3 deficiencies or in any way relieve the SELLER of his obligations under the Contract.
- EQUIPMENT rejected by the PURCHASER/ CONSULTANT shall be removed by the Seller at his cost within 14 days of notice after repaying the amounts received against the SUPPLY. 23.4 The PURCHASER shall in no way be responsible for any deterioration or damage to the EQUIPMENT under any circumstances whatsoever.
- 23.5 In case of rejection of EQUIPMENT, PURCHASER shall have the right to recover the amounts, if any, from any of CONTRACTOR'S invoices pending with PURCHASER or by alternative method(s).

24.1 Termination for Default 24.1.1

In the event that the SELLER commits a breach of the provisions of this CONTRACT and/or fails to perform any of its obligations under this CONTRACT including but not limited to failing to deliver any or all of the GOODS within the time period (s) specified in the CONTRACT, the SELLER shall be said to have committed a default of the provisions of this CONTRACT. In the event that such default is not cured by the SELLER within 30 days of a notice issued by the PURCHASER of such default, the PURCHASER shall be entitled to terminate this CONTRACT forthwith. The PURCHASER may, at its sole discretion, extend the time for curing the default by the SELLER beyond 30 days. The PURCHASER may, without prejudice to any other remedy for breach of CONTRACT, by written notice of default sent to the SELLER, terminate the CONTRACT in whole or in part:

A) If the SELLER fails to deliver any or all of the GOODS within the time period(s) specified in the CONTRACT; or

B) If the SELLER fails to perform any other obligation(s) under the CONTRACT, and

- C) If the SELLER, in either of the above circumstances, does not cure his failure within a period of 30 days (or such longer period as the PURCHASER may authorize in writing) after receipt of the default notice from the PURCHASER
- 24.1.2. In the event the PURCHASER terminates the CONTRACT in whole or in part, pursuant to Article 24.1.1the PURCHASER may procure, upon such terms and in such manner as it deems appropriate, goods similar to those undelivered and the SELLER shall be liable to the PURCHASER for any excess costs for such similar GOODS. However, the SELLER shall continue performance of the CONTRACT to the extent not terminated.
- 24.1.3 In case of termination of CONTRACT herein set forth (under clause 24) except under conditions of Force Majeure and termination after expiry of contract, the VENDOR shall be blacklisted/ Debarred as per prevailing guidelines.

24.2 Termination for Insolvency

24.2.1 The PURCHASER, may at any time, terminate the CONTRACT by giving written notice to the SELLER, without compensation to the SELLER, if the SELLER becomes bankrupt or otherwise insolvent, provided that such termination will not prejudice or affect any right of action or remedy which has accrued or will accrue thereafter to the PURCHASER.

Force Majeure 25.0

Force majeure is an event beyond the control of SELLER and not involving the SELLER's fault or negligence and which is not foreseeable, which prevents the SELLER from fulfilling its 25.1 obligations under the CONTRACT. Force Majeure shall mean and be limited to the following: a) War/hostilities

- b) Riot or Civil commotion c) Earthquake, flood, tempest, lightening or other natural physical disaster.
- d) Restrictions imposed by the Government or other Statutory bodies which prevent or delay the execution of the Contract by the SELLER.
- The decision as to whether an event constitutes force majeure shall rest with PURCHASER which shall be final and binding.

- 25.2 If there is delay in performance or other failures by the SELLER to perform obligations under the contract due to event of a Force Majeure, the SELLER shall not be held responsible for such delays/failures.
- 25.3 If a Force Majeure situation arises, the SELLER shall promptly notify the purchaser in writing of such conditions and the cause thereof within fifteen days of occurrence of such event.

 Unless otherwise directed by the Purchaser in writing, the SELLER shall continue to perform its obligations under the contract as far as reasonable/practical and shall seek all reasonable alternative means for performance not prevented by the Force Majeure event.
- 25.4 If the performance in whole or in part or any obligation under this contract is prevented or delayed by any reason of Force Majeure for a period of exceeding 120 days, PURCHASER may at its option terminate the contract without any financial repercussion on either side.

26.0 Resolution of Disputes/ Arbitration:

- The PURCHASER and the SELLER shall make every effort to resolve amicably by direct informal, good faith negotiations any disagreement or dispute arising between them under or in connection with the contract.
- 26.2 If, after thirty days from the commencement of such informal negotiations, the PURCHASER and the SELLER have been unable to resolve the disagreement or dispute, the same be referred for resolution to the formal mechanism as specified hereunder.
- 26.3 Legal Construction: The Contract shall in all respects be construed and operated as an Indian Contract and in accordance with Indian Laws as in force for the time being.

Arbitration:

a) All disputes and differences of any kind whatsoever arising out of or in connection with the contract or carrying out of the works (whether during the course of works or after their completion and whether before or after termination, abandonment or breach of contract) shall be referred to and settled by the person authorized and notified in writing by PURCHASER who shall state his decision in writing. Such a decision may be in the form of a final certificate or otherwise and shall be made within a period of 30 days from the date of receipt of such reference to them.

b) If the SELLER is dissatisfied with the decision of such authorized person, then he may within 30 days of receipt of such decision send a written appeal to PURCHASER, represented by the Chairman and Managing Director at the registered office, Mumbai for the same to be referred to Arbitration by a Sole Arbitrator to be appointed by mutual consent and after due approval of CMD, IREL. The Arbitration proceedings shall be conducted as per the provisions of the Arbitration and Conciliation Act, 1996. It is made clear that this Arbitration Clause shall be applicable to any and all disputes and differences between the Parties arising out of and/or relating to this CONTRACT and the Parties shall be bound to refer the same to arbitration in accordance with the procedure contemplated herein

- c) If the period of 30 days under Clause (b) has expired at any stage, stipulated in the preceding paras without any response from the SELLER before such expiry, the SELLER is deemed to have communicated his satisfaction to the decision of PURCHASER at the relevant stage and all his rights of further appeal or as the case may be and adjudication are deemed to have been waived once and for all.
- d) The seat of arbitration will be at Mumbaiand language thereof shall be English.
- e) Notwithstanding the invocation, commencement and/or pendency any dispute resolution proceedings under this Clause 26.0 including arbitration under Clause 26.4, the SELLER shall continue to be bound by the provisions of the CONTRACT, if not terminated by the EMPLOYER, and shall be obligated to discharge its obligations under the CONTRACT including continuation of the WORK under the CONTRACT.
- f) The Bidder shall not in any way delay or default or cause to delay or default the carrying out of the works by reason of the fact that any matter has been agreed to be referred for dispute resolution under Clause 26.0 including Arbitration under Clause 26.4.
- 26.5 Jurisdiction The courts at Kochi only shall, subject to Arbitration Clause, have exclusive jurisdiction to deal with and decide any matter arising out of this contract.

27. Taxes & Duties

- 27.1 A foreign Seller shall be entirely responsible for all taxes, stamp duties, licence fees, and other such levies imposed outside the PURCHASER's country.
- A domestic Seller shall be entirely responsible for all taxes, duties, licence fees etc. incurred until the delivery of the contracted goods to the PURCHASER. However, GST on finished products shall be reimbursed by PURCHASER. Seller need to prepare E-Invoice under GST and submit monthly/ quarterly GST Return as per GST Rules. Failure to submit GST Return on time as per GST Rules may result into deduction of GST while processing suppliers' Bill.
- 27.3 Customs duty payable in India for imported goods ordered by PÜRCHASER on foreign Seller shall be borne and paid by PÜRCHASER.
- 27.4 Any income tax payable in respect of supervisory services rendered by foreign Seller under the Contract shall be as per the Indian Income Tax Act and shall be borne by SELLER. It is upto the bidder/seller to ascertain the amount of these taxes and to include them in his bid price.

28. Permits & Certificates 28.1 SELLER shall procure

SELLER shall procure, at his expense, all necessary permits, certificates and licences required by virtue of all applicable laws, regulations, ordinances and other rules in effect at the place where any of the work is to be performed, and SELLER further agrees to hold PURCHASER and/or CONSULTANT harmless from liability or penalty which might be imposed by reason of any asserted or established violation of such laws, regulations, ordinances or other rules. PURCHASER will provide necessary permits for SELLER's personnel to undertake any work in India in connection with Contract.

Fall Clause

29.2

- 29.1 The price charged for the materials supplied under the order by the supplier shall in no event exceed the lowest price at which the supplier or his agent/principal/dealer, as the case may be, sells the materials of identical description to any Persons/Organizations including the Purchaser or any Department of the Central Govt. or any Department of a State Govt. or any Statutory Undertaking of the Central or State Govt. as the case may be, during the currency of the order.
 - If at any time during the said period, the supplier or his agent/principal/dealer, as the case may be, reduces the sale price, sells or offers to sell such materials to any persons/organizations including the Purchaser or any Deptt. Of Central Govt. or State Govt. as the case may be, at a price lower than the price chargeable under the order, he shall forthwith notify such reduction or sale or offer of sale to the Purchase Authority who has issued this order and the price payable under the order for the materials supplied after the date of coming into force of such reduction or sale or offer of sale shall stand correspondingly reduced. The above stipulation will, however, not apply to:
 - a) Exports by the Contractor/Supplier or
 - b) Sale of goods as original equipment at prices lower than the prices charged for normal replacement
 - c) Sale of goods such as drugs which have expiry dates.
- 29.3 The supplier shall furnish the following certificate to the concerned Paying Authority along with each bill for payment for supplies made against this order:- "I/We certify that there has been no reduction in sale price of the items/goods/materials of description identical to those supplied to the PURCHASER under the order herein and such items/goods/materials have not been offered/sold by me/us to any person/organizations including the Purchaser or any Deptt. of Central Govt. or any Deptt. of State Govt. or any Statutory Undertaking of the Central or State Govt. as the case may be upto the date of bill/during the currency of the order whichever is later, at a price lower than the price charged to the PURCHASER under the order". Such a certificate shall be obtained, except for quantity of items/goods/materials categories under sub-clause (a), (b) & (c) of sub-para 29.2 above, of which details shall be furnished by the supplier.

30.0 Limitation of Liability

Notwithstanding anything contrary contained herein, the aggregate total liability of Seller under the Agreement or otherwise shall be limited to 100% of Agreement / Order price. However, neither party shall be liable to the other party for any indirect and consequential damages, loss of profits or loss of production.

31.0 Method of blacklisting vendors

- Any failure by the vendor (SELLER) to supply/execute the contract as per order may result in blacklisting of vendors by the PURCHASER through it competent authority. The blacklisted vendor shall not be considered for a minimum period of one year from the date of blacklisting.
- Further the competent authority may blacklist the bidder, if the bidder changes bid (either techno-commercial and / or price) or withdraw his bid after receipt of the same and during the validity period of bid.
- 31.3 Further, the vendor (SELLER) shall be banned from doing any business with the PURCHASER, indefinitely or for such period as may be determined by the PURCHASER at it sole discretion, in case of:
 - a. If security considerations including question of loyalty to the state so warrant.
 - b. If the proprietor, partner, Director's representative of the SELLER is convicted by a court of law following prosecution for any offences including offences relating to business or commercial dealings.
 - c. If there is strong justification, as determined by the PURCHASER at its sole discretion, for believing that the proprietor or employee or representative of the SELLER has been guilty of malpractice such as bribery, corruption, fraud, substitution of tenders, interpolation, misrepresentation, evasion or habitual default in payment of any tax levied by law, etc.

- An order for ban/ suspension passed for a certain specified period shall be deemed to have been automatically revoked on expiry of that specified period and it will not be necessary to issue a specific formal order of revocation, except that an order of suspension/ban passed on account of doubtful loyalty or security consideration shall continue to remain in force until it is specifically revoked in writing.
- 31.5 An order of ban on grounds of conviction by Court of Law may be revoked if, in respect of the same facts, the SELLER has been wholly acquitted by a court of law.

32.0 Secre

The SELLER shall not at any time during the pendency of the contract or thereafter disclose any information furnished to them by PURCHASER or any drawings, designs, reports and other documents and information prepared by the Contractor for this contract, without the prior written approval of PURCHASER except in so far as such disclosure is necessary for the performance of the Contractor's work and service hereunder.

General

34.0

- 33.1 In the event that terms and conditions stipulated in the General Conditions of Contract should deviate from terms and conditions stipulated in the Contract, the latter shall prevail. Any specific Terms / Conditions mentioned in PO will supersede this GCC.
- 33.2 Losses due to non-compliance of Instructions: Losses or damages occurring to the PURCHASER owing to the SELLER's failure to adhere to any of the instructions given by the PURCHASER/CONSULTANT in connection with the contract execution shall be recoverable from the SELLER.
- 33.3 Recovery of sums due: All costs, damages or expenses which the PURCHASER/CONSULTANT may have paid, for which under the contract, SELLER is liable, may be recovered by the PURCHASER (he is hereby irrevocably authorized to do so) from any money due to or becoming due to the SELLER under this Contract or other Contracts and/or may be recovered by action at law or otherwise. If the same due to the SELLER be not sufficient to recover the recoverable amount, the SELLER shall pay to the PURCHASER, on demand, the balance amount.
- action at law or otherwise. If the same due to the SELLER be not sufficient to recover the recoverable amount, the SELLER shall pay to the PURCHASER, on demand, the balance amount.

 Payments, etc. not to affect rights of the PURCHASER: No sum paid on account by the PURCHASER nor any extension of the date for completion granted by the PURCHASER shall affect or prejudice the rights of the PURCHASER against the SELLER or relieve the SELLER of his obligation for the due fulfillment of the CONTRACT.
- 33.5 Cut-off Dates: No claims or correspondence on this Contract shall be entertained by the PURCHASER/Consultant after 90 days after expiry of the performance guarantee (from the date of final extension, if any).
- 33.6 Paragraph heading: The paragraph heading in these conditions shall not affect the construction thereof

RISK PURCHASE CLAUSE

After award of the contract, if the tenderer fails to execute the work as per tender or at any time repudiates the order, IREL (India) Limited has the right to forfeit and invoke the security deposit and execute the order from other agencies at the risk and cost of the tenderer. The cost difference between the alternative arrangements and total contract value will be recovered from the tenderer along with other incidental charges. In case of execution of order through alternative sources and if price is lower, no benefit on this account will be passed on to the tenderer.

"ETHICS IN TENDERING & OTHER BUSINESS DEALINGS"

Dear Sir,

IREL (India) LIMITED, a Government of India undertaking under the administrative control of Department of Atomic Energy is doing its business as per the rules and regulation of the Public Sector Undertaking and other statutory agencies. The business is done in an ethical, rational & impartial manner with good corporate governance.

In our endeavour to be more transparent in our dealings and to support our ideology all Vendors, Customers and Business Partners are requested not to provide any gift and / or inducement to any of our employees for securing / being granted favour in dealings with our Company. In assurance of your commitment to the aforesaid, it will be highly appreciated if you fill up, sign and abide by the attached undertakings.

Report of any gifts and / or inducements sought by any employee of the company should be immediately reported to any one of the following:

Chairman & Managing Director	Chief Vigilance Officer
IREL (India) Limited	IREL (India) Limited
1207, V.S. Marg, Prabhadevi	1207, V.S. Marg, Prabhadevi
Mumbai 400 028.	Mumbai 400 028.
Ph: 022-24225778	Ph: 022-24221068
E-mail: cmd@irel.co.in	E-mail: cvo@irel.co.in

We assure you that complaints if any made by you on the subject will be kept confidential and fair investigation will be conducted and appropriate action will be taken. Similarly, we accept your commitment to the undertaking and its violation will have consequences as per prevailing rule of the Company.

Thanking you,
For IREL (India) Limited
Name
Designation
Date:

आईआरईएल (इंडिया) लिमिटेड IREL (India) Limited

रेअरअर्थ्सप्रभाग Rare Earths Division

TECHNICAL DEVIATION STATEMENT FORM

How form	rever in case of any deviation, 		at of the bid. No deviation is expected. In the technical deviation statement of the Tender specifications.
	CLAUSE	DEVIATION (Including Justification)	REMARKS
Sign	ature & Seal of the Tenderer		
Date	d:		
Note	<u>:</u>		
		, the statement should be returne not endorsed, it will be considered as	ed duly signed with an endorsement s 'No deviation
b. T	Γhe technical specifications fu	urnished in the Tender document sh	all prevail over those of any other

document forming a part of our Tender, except only to the extent of deviations furnished in the statement.

TENDER NO: Name of Work:

UNDERTAKING TO BE SUBMITTED BY BIDDERS FOR ADOPTING ETHICAL PRACTICES

Date:

IREL (India) Limited
1207, V.S. Marg, Prabhadevi
Mumbai 400 028.
I / We
(now onwards to be referred as Company).
I / We agree and undertake:
Not to provide any gift and / or inducement to any employee of the Company in connection with securing /
being granted favour (s) in my / our dealings with the Corporate office of the company and / or its any field units namely MK, Chavara, OSCOM, RED & IRERC.
To immediately report any gift and / or inducement sought by any employee of the Company granting favour(s) to me / us in my / our dealings with the Company and / or its field units.
Signature
Name
Title
Name of the Company and Address (with Seal)

IREL Bank Details

Name of Bank	State Bank of India, Udyogamandal Branch, Udyogamandal
Account Type	Current Account
Name of Account holder	IREL(India) Limited
Account Number	57017844321
IFSC Code	SBIN0070158
MICR Code	682002926

Details of payment to be uploaded as attachment during tender submission and the same shall be e-mailed to: purchase-red@irel.co.in

Bidders (Pl. Provide the details):			
1. Name of the Bank :			
2. Account Type :			
3. Account Number :			
4. IFSC Code :			

		VENDOR UP-DAT	ION DETAILS	S		
SI#	Organisation Details					
1	Name					
2	`	address / Office Address / dress / Factory Address)				
	Building / Hou	se Number				
	Area/Street Na	ame				
	City					
	State					
	Pin Code					
			Ph No:			
	Contact Details	3	Mobile No.			
			Email:			
3	Vendor Type (Domestic / Import)				
4	Company/ Or Liability Partr Proprietorship,	Information (Private Limited ne Person Company/ Limited nership/ Partnership Firm // Co-Operative Society/ Trust/				
5	Trader/ Distr Partner/ India	siness (Manufacturer/ Dealer/ ributor/ Stockiest/ Channel an Sales Office/ Subsidiary of eign supplier/ Indian Agents/ er)				
6	Whether Approved By Any Central Government/State Government/ Public Sector Undertakings / Reputed Private Organisations for similar items, for which registration is sought.					
7	PAN No.	Copy of the same may be				
8	GSTIN ID	attached.				
9	-	of P&L for the last three FY				
10	Valid MSE Udyam registration certificate, if any.		Yes	No		
11	MSE ownership	o details.		1		
12	ISO Certification if any					
13	Registered in C	GEM Portal	Yes	No		
14	MSE to confir TReDs platform	m if they are registered from	Yes	No		
15	1 1	ier/Service Provider is a Start- If yes provide the details.				
16	Whether sh Litigation/Arb years?	e Company is under itration cases during last 5	Yes	No		

17	Whether company has been delisted/debarred from business with any PSU in India, and the reasons thereof?	
18	Bank D	Details
	Name of bank:	
	Name of Bank Branch:	
	City/Place:	
	Account Number:	
	Account Type:	
	IFSC Code:	
	MICR Code:	
	Swift Code	
	Self-attested or Bank attested Bank details on Company letterhead or cancelled cheque	
	Current year Solvency Certificate	

PROFORMA FOR BANK GUARANTEE FOR EARNEST MONEY DEPOSIT

In accordance with Notice Inviting Tender (NIT) No	Dated	for the work of
(herein after referred to as "the said Works") for Rs	(Rupees_	only), under RED
unit of M/s IREL (India) Limited, a company incorporated	under Indian Compa	nies Act, having its registered
office at Plot No.1207, ECIL building, Opp. to Siddhiving	_	
Mumbai - 400028, India (herein after referred to		
[Herein after referred to as Co	, , =	
tender and a Bank Guarantee for the sum of Rs.		for a period of days
(in words) is required to be submitted by the Bidder toward	ls the Bid Security.	
We theBank (hereinafter	called the said Bank)	do hereby undertake to pay to
IREL, the sum of Rs (Rupees		
enter into an agreement of contract on intimation of accept	ance of his tender and	or to commence the contract
works and/or failure to deposit the security deposit w	ithin the stipulated p	period as per the terms and
conditions relating to and/or governing the contract and/or	-	
also agree that any such demand made on the Bank shall be	_	
by the Bank under this Guarantee. We also agree that	0 ,	- ,
litigation in respect of or arising from the said contract and,	-	
stated by IREL including the question as to the tenability Money being the Bank Guarantee herein, we shall forthw		· ·
made as aforesaid.	itii pay tile said airiou	int to IKEL on demand being
made as aforesafd.		
We Bank further agree that the guaran	tee herein contained s	shall remain in full force and
effect during the period that would be taken for enterin	ig into an Agreement	of contract and that it shall
continue to be enforceable till all the dues of the IREL und		
have been fully paid and its claims satisfied or discharged or		
the NIT have been fully and properly carried out by	the said tenderer and	daccordingly discharges the
guarantee.		
We Bank further agree with the IRE	EL that the IREL shall h	nave the fullest liberty without
our consent and without affecting in any manner our ob	oligations hereunder t	o vary any of the terms and
conditions of the NIT and/or terms and conditions govern	ning the contract or to	extend the time of validity of
the offer from the said tenderer from time to time or to post	-	
exercisable by the IREL against the said tenderer and to fo	•	
the NIT and we shall not be relieved from our liability here	-	
being granted to the said tenderer or for any forbearance		-
indulgence by the IREL to the said tenderer or by any such		
relating to surety/guarantee would but for this provision has	ave effect of so relieving	ıg us.
WeBank do hereby further agree	ee that any change in	the Constitution of the said
tenderer or the Bank will not affect the validity of this guara		
David landles and Januari and Land	ovolto this comment	leuring its gramor
We Bank lastly undertake not to re	evoke trus guarantee d	uring its currency except with
the previous consent of the IREL in writing.		

Notwithsta	anding anything	g to the contrary	contained	herein befor	e:					
(i) (ii) (iii)	Rs (ii) This Bank (iii) The bank only if IREL	y under this	(in words) be valid u he guaran claim or	ptoteed amount demand on	or any p	oart the	unless ex ereof und	tended er this	on demand bank Guarai	ntee
	Dated the	day of _ Bank		_ 20 —						
				(Signatu	re with 1	name i	n Block le	etters w	ith designat	ion,
				Attorne	ey as per	powe	r of Attorn	ney No	dt)
								Bank	's Common	seal





Bid Number/बोली क्रमांक (बिड संख्या)[:] GEM/2025/B/6085530 Dated/दिनांक : 25-03-2025

Bid Document/ बिड दस्तावेज़

Bid Details/बिड विवरण			
Bid End Date/Time/बिड बंद होने की तारीख/समय	15-04-2025 13:00:00		
Bid Opening Date/Time/बिंड खुलने की तारीख/समय	15-04-2025 13:30:00		
Bid Offer Validity (From End Date)/बिड पेशकश वैधता (बंद होने की तारीख से)	90 (Days)		
Ministry/State Name/मंत्रालय/राज्य का नाम	Pmo		
Department Name/विभाग का नाम	Department Of Atomic Energy		
Organisation Name/संगठन का नाम	Irel (india) Limited		
Office Name/कार्यालय का नाम	Но		
क्रेता ईमेल/Buyer Email	bycon1.irel.kerala@gembuyer.in		
Total Quantity/कुल मात्रा	4		
ltem Category/मद केटेगरी	Supply, installation, commissioning, testing & training of HP Ge detector based gamma spectrometer a , Gamma spectrometer AMC for First year , Gamma spectrometer AMC for second year , Gamma spectrometer AMC for third year		
GeMARPTS में खोजी गई स्ट्रिंग / Searched Strings used in GeMARPTS	Gamma spectrometer AMC for third year		
GeMARPTS में खोजा गया परिणाम / Searched Result generated in GeMARPTS	Category not available on GeM for the text string uploaded by the buyer		
अधिसूचना के लिए चयनित प्रासंगिक श्रेणियाँ / Relevant Categories selected for notification	Optical Emmision Spectrometer		
Minimum Average Annual Turnover of the bidder (For 3 Years)/बिडर का न्यूनतम औसत वार्षिक टर्नओवर (3 वर्षों का)	58 Lakh (s)		
Years of Past Experience Required for same/similar service/उन्हीं/समान सेवाओं के लिए अपेक्षित विगत अनुभव के वर्ष	3 Year (s)		
MSE Exemption for Years Of Experience/अनुभव के वर्षों से एमएसई छ्ट/ and Turnover/टर्नओवर के लिए एमएसई को छूट प्राप्त है	Yes		

Bid Details/बिड विवरण			
Startup Exemption for Years Of Experience/अनुभव के वर्षों से स्टार्टअप छूट/ and Turnover/ टर्नओवर के लिए स्टार्टअप को छूट प्राप्त है	Yes		
Document required from seller/विक्रेता से मांगे गए दस्तावेज़	Experience Criteria, Bidder Turnover, Certificate (Requested in ATC), OEM Authorization Certificate, Additional Doc 1 (Requested in ATC), Additional Doc 2 (Requested in ATC), Additional Doc 3 (Requested in ATC), Compliance of BoQ specification and supporting document *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer		
Do you want to show documents uploaded by bidders to all bidders participated in bid?/	No		
Bid to RA enabled/बिंड से रिवर्स नीलामी सक्रिय किया	No		
Type of Bid/बिंड का प्रकार	Two Packet Bid		
Primary product category	Supply, installation, commissioning, testing & training of HP Ge detector based gamma spectrometer a		
Time allowed for Technical Clarifications during technical evaluation/तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय	2 Days		
Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM)	No		
Payment Timelines	Payments shall be made to the Seller within 30 days of issue of consignee receipt-cum-acceptance certificate (CRAC) and on-line submission of bills (This is in supersession of 10 days time as provided in clause 12 of GeM GTC)		
Evaluation Method/मूल्यांकन पद्धति	Total value wise evaluation		
Financial Document Required/वितीय दस्तावेज की आवश्यकता है।	Yes		
Arbitration Clause	No		
Mediation Clause	No		

EMD Detail/ईएमडी विवरण

Advisory Bank/एडवाईजरी बैंक	State Bank of India	
EMD Amount/ईएमडी राशि	328800	

ePBG Detail/ईपीबीजी विवरण

Advisory Bank/एडवाइजरी बैंक	State Bank of India	
ePBG Percentage(%)/ईपीबीजी प्रतिशत (%)	5.00	
Duration of ePBG required (Months)/ईपीबीजी की अपेक्षित अवधि (महीने).	26	

- (a). EMD EXEMPTION: The bidder seeking EMD exemption, must submit the valid supporting document for the relevant category as per GeM GTC with the bid. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded from the purview of this Policy,/जेम की शर्तों के अनुसार ईएमडी छूट के इच्छुक बिडर को संबंधित केटेगरी के लिए बिड के साथ वैध समर्थित दस्तावेज़ प्रस्तुत करने है। एमएसई केटेगरी के अंतर्गत केवल वस्तुओं के लिए विनिर्माता तथा सेवाओं के लिए सेवा प्रदाता ईएमडी से छूट के पात्र हैं। व्यापारियों को इस नीति के दायरे से बाहर रखा गया है।
- (b). EMD & Performance security should be in favour of Beneficiary, wherever it is applicable./ईएमडी और संपादन जमानत राशि, जहां यह लागू होती है, लाभार्थी के पक्ष में होनी चाहिए।

Beneficiary/लाभार्थी :

DGM (A&E)

IREL (India) Limited, RED, Udyogammandal (Kamalam.p)

MII Purchase Preference/एमआईआई खरीद वरीयता

(()	
MII Purchase Preference/एमआईआई खरीद वरीयता	Yes

MSE Purchase Preference/एमएसई खरीद वरीयता

Yes

- 1. If the bidder is a Micro or Small Enterprise as per latest definitions under MSME rules, the bidder shall be exempted from the requirement of "Bidder Turnover" criteria and "Experience Criteria" subject to meeting of quality and technical specifications. If the bidder is OEM of the offered products, it would be exempted from the "OEM Average Turnover" criteria also subject to meeting of quality and technical specifications. In case any bidder is seeking exemption from Turnover / Experience Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer.
- 2. If the bidder is a Startup, the bidder shall be exempted from the requirement of "Bidder Turnover" criteria and "Experience Criteria" subject to their meeting of quality and technical specifications. If the bidder is OEM of the offered products, it would be exempted from the "OEM Average Turnover" criteria also subject to meeting of quality and technical specifications. In case any bidder is seeking exemption from Turnover / Experience Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer.
- 3. The minimum average annual financial turnover of the bidder during the last three years, ending on 31st March of the previous financial year, should be as indicated above in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the bidder is less than 3-year-old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.
- 4. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU for number of Financial years as indicated above in the bid document before the bid opening date. Copies of relevant contracts to be submitted along with bid in support of having supplied some quantity during each of the Financial year. In case of bunch bids, the category of primary product

having highest value should meet this criterion.

5. Preference to Make In India products (For bids < 200 Crore):Preference shall be given to Class 1 local supplier as defined in public procurement (Preference to Make in India), Order 2017 as amended from time to time and its subsequent Orders/Notifications issued by concerned Nodal Ministry for specific Goods/Products. The minimum local content to qualify as a Class 1 local supplier is denoted in the bid document. If the bidder wants to avail the Purchase preference, the bidder must upload a certificate from the OEM regarding the percentage of the local content and the details of locations at which the local value addition is made along with their bid, failing which no purchase preference shall be granted. In case the bid value is more than Rs 10 Crore, the declaration relating to percentage of local content shall be certified by the statutory auditor or cost auditor, if the OEM is a company and by a practicing cost accountant or a chartered accountant for OEMs other than companies as per the Public Procurement (preference to Make-in -India) order 2017 dated 04.06.2020. Only Class-I and Class-II Local suppliers as per MII order dated 4.6.2020 will be eligible to bid. Non - Local suppliers as per MII order dated 04.06.2020 are not eligible to participate. However, eligible micro and small enterprises will be allowed to participate .The buyers are advised to refer the OM No.F.1/4/2021-PPD dated 18.05.2023.

OM_No.1_4_2021_PPD_dated_18.05.2023 for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017.

6. Purchase preference will be given to MSEs having valid Udyam Registration and whose credentials are validated online through Udyam Registration portal as defined in Public Procurement Policy for Micro and Small Enterprises (MSEs) Order, 2012 dated 23.03.2012 issued by Ministry of Micro, Small and Medium Enterprises and its subsequent Orders/Notifications issued by concerned Ministry. If the bidder wants to avail themselves of the Purchase preference, the bidder must be the manufacturer / OEM of the offered product on GeM. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises and hence resellers offering products manufactured by some other OEM are not eligible for any purchase preference. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service and Buyer will decide eligibility for purchase preference based on documentary evidence submitted, while evaluating the bid. If L-1 is not an MSE and MSE Seller (s) has / have quoted price within L-1+ 15% (Selected by Buyer) of margin of purchase preference /price band defined in relevant policy, such MSE Seller shall be given opportunity to match L-1 price and contract will be awarded for 25% (selected by Buyer) percentage of total quantity. The buyers are advised to refer the OM No. F.1/4/2021-PPD dated 18.05.2023 OM No.1 4 2021 PPD dated 18.05.2023 for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017. Benefits of MSE will be allowed only if seller is validated on-line in GeM profile as well as validated and approved by Buyer after evaluation of documents submitted.

Pre Bid Detail(s)

Pre-Bid Date and Time	Pre-Bid Venue
04-04-2025 11:00:00	IREL (India) Limited, RED, Udyogamandal

Supply, Installation, Commissioning, Testing & Training Of HP Ge Detector Based Gamma Spectrometer A (1 pieces)

(Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively/क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक)

Technical Specifications/तकनीकी विशिष्टियाँ

Buyer Specification Document/क्रेता विशिष्टि दस्तावेज़	Download
---	----------

Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Anil Kumar.V.A	683501,IREL (India) Limited; UDYOGMANDAL,Ernakulam.	1	270

Gamma Spectrometer AMC For First Year (1 pieces)

(Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively/क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक)

Technical Specifications/तकनीकी विशिष्टियाँ

1 1			_
	Buyer Specification Document/क्रेता विशिष्टि दस्तावेज़	<u>Download</u>	

Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती / रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Anil Kumar.V.A	683501,IREL (India) Limited; UDYOGMANDAL,Ernakulam.	1	365

Gamma Spectrometer AMC For Second Year (1 pieces)

(Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively/क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक)

Technical Specifications/तकनीकी विशिष्टियाँ

Buyer Specification Document/क्रेता विशिष्टि दस्तावेज़	<u>Download</u>
---	-----------------

Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Anil Kumar.V.A	683501,IREL (India) Limited; UDYOGMANDAL,Ernakulam.	1	365

Gamma Spectrometer AMC For Third Year (1 pieces)

(Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively/क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक)

Technical Specifications/तकनीकी विशिष्टियाँ

Buyer Specification Document/क्रेता विशिष्टि दस्तावेज़	Download
Bocamengs (ii 141(ii) 4((ii49)	

Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Anil Kumar.V.A	683501,IREL (India) Limited; UDYOGMANDAL,Ernakulam.	1	365

Buyer Added Bid Specific Terms and Conditions/क्रेता द्वारा जोड़ी गई बिड की विशेष शर्तें

1. Generic

OPTION CLAUSE: The Purchaser reserves the right to increase or decrease the quantity to be ordered up to 25 percent of bid quantity at the time of placement of contract. The purchaser also reserves the right to increase the ordered quantity up to 25% of the contracted quantity during the currency of the contract at the contracted rates. The delivery period of quantity shall commence from the last date of original delivery order and in cases where option clause is exercised during the extended delivery period the additional time shall commence from the last date of extended delivery period. The additional delivery time shall be (Increased quantity ÷ Original quantity) × Original delivery period (in days), subject to minimum of 30 days. If the original delivery period is less than 30 days, the additional time equals the original delivery period. The Purchaser may extend this calculated delivery duration up to the original delivery period while exercising the option clause. Bidders must comply with these terms.

2. Generic

Bidder shall submit the following documents along with their bid for Vendor Code Creation:

a. Copy of PAN Card.

- b. Copy of GSTIN.
- c. Copy of Cancelled Cheque.
- d. Copy of EFT Mandate duly certified by Bank.

Disclaimer/अस्वीकरण

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent Authority in Buyer Organization, whereby Buyer organization is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any eccentricity / restriction arising in the bidding process due to these ATCs and due to modification of technical specifications and / or terms and conditions governing the bid. If any clause(s) is / are incorporated by the Buyer regarding following, the bid and resultant contracts shall be treated as null and void and such bids may be cancelled by GeM at any stage of bidding process without any notice:-

- 1. Definition of Class I and Class II suppliers in the bid not in line with the extant Order / Office Memorandum issued by DPIIT in this regard.
- 2. Seeking EMD submission from bidder(s), including via Additional Terms & Conditions, in contravention to exemption provided to such sellers under GeM GTC.
- 3. Publishing Custom / BOQ bids for items for which regular GeM categories are available without any Category item bunched with it.
- 4. Creating BoQ bid for single item.
- 5. Mentioning specific Brand or Make or Model or Manufacturer or Dealer name.
- 6. Mandating submission of documents in physical form as a pre-requisite to qualify bidders.
- 7. Floating / creation of work contracts as Custom Bids in Services.
- 8. Seeking sample with bid or approval of samples during bid evaluation process. (However, in bids for <u>attached categories</u>, trials are allowed as per approved procurement policy of the buyer nodal Ministries)
- 9. Mandating foreign / international certifications even in case of existence of Indian Standards without specifying equivalent Indian Certification / standards.
- 10. Seeking experience from specific organization / department / institute only or from foreign / export experience.
- 11. Creating bid for items from irrelevant categories.
- 12. Incorporating any clause against the MSME policy and Preference to Make in India Policy.
- 13. Reference of conditions published on any external site or reference to external documents/clauses.
- 14. Asking for any Tender fee / Bid Participation fee / Auction fee in case of Bids / Forward Auction, as the case may be.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller within 4 days of bid publication on GeM. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

All GeM Sellers / Service Providers are mandated to ensure compliance with all the applicable laws / acts / rules including but not limited to all Labour Laws such as The Minimum Wages Act, 1948, The Payment of Wages Act, 1936, The Payment of Bonus Act, 1965, The Equal Remuneration Act, 1976, The Payment of Gratuity Act, 1972 etc. Any non-compliance will be treated as breach of contract and Buyer may take suitable actions as per GeM Contract.

This Bid is also governed by the General Terms and Conditions/ यह बिड सामान्य शर्तों के अंतर्गत भी शासित है

In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws./जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध में भारत के साथ भूमि सीमा साझा करने वाले देश के लिए तभी पात्र होगा

जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो।बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्रवाई का आधार होगा।

---Thank You/धन्यवाद---